


LB122089

SFAM01.1

INSTRUMENT ID: CV2

 μ Abs.:

43441



Linear

A= 0.0000e+000

B= 2.3115e-004

C= -1.5419e-002

Rho= 0.9999648

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.018	0.018	146	0.000	146					17
0.05	0.050					185	0				-2
0.20	0.200	0.234	0.034	1078	0.0 %	1078					-1
2.50	2.500	2.459	-0.041	10705	0.0 %	10705					0
5.00	5.000	4.957	-0.043	21510	0.0 %	21510					0
7.50	7.500	7.506	0.006	32541	0.0 %	32541					0
10.0	10.000	10.026	0.026	43441	0.0 %	43441					0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	146	-		0 SFAM01.1	PPB	9/20/2022 14:59	Std
	0.2 S01	1078	-		0.2 SFAM01.1	PPB	9/20/2022 15:01	Std
	2.5 S02	10705	-		2.5 SFAM01.1	PPB	9/20/2022 15:03	Std
	5 S03	21510	-		5 SFAM01.1	PPB	9/20/2022 15:05	Std
	7.5 S04	32541	-		7.5 SFAM01.1	PPB	9/20/2022 15:08	Std
	10 S05	43441	-		10 SFAM01.1	PPB	9/20/2022 15:10	Std
ICV037	ICV037	17081	3.9328 -		SFAM01.1	PPB	9/20/2022 15:13	SMPL
ICB037	ICB037	72	0.0012 -		SFAM01.1	PPB	9/20/2022 15:15	SMPL
CCV035	CCV035	21227	4.8912 -		SFAM01.1	PPB	9/20/2022 15:17	SMPL
CCB035	CCB035	49	-0.0041 -		SFAM01.1	PPB	9/20/2022 15:20	SMPL
PB147772BL	PBW772	94	0.0063 -		SFAM01.1	PPB	9/20/2022 15:24	SMPL
N4015-01	MDL-WATER	272	0.0475 -		SFAM01.1	PPB	9/20/2022 15:27	SMPL
PB147773BL	PBW773	86	0.0045 -		SFAM01.1	PPB	9/20/2022 15:29	SMPL
N4543-01	YBF57	1406	0.3096 -		SFAM01.1	PPB	9/20/2022 15:31	SMPL
N4543-02	YBF57D	1391	0.3061 -		SFAM01.1	PPB	9/20/2022 15:33	SMPL
N4543-03	YBF57S	6063	1.386 -		SFAM01.1	PPB	9/20/2022 15:36	SMPL
N4543-04	YBF59	786	0.1663 -		SFAM01.1	PPB	9/20/2022 15:38	SMPL
N4543-05	YBFB7	1353	0.2973 -		SFAM01.1	PPB	9/20/2022 15:40	SMPL
N4543-06	YBFB9	1279	0.2802 -		SFAM01.1	PPB	9/20/2022 15:43	SMPL
N4543-07	YBFC1	1170	0.255 -		SFAM01.1	PPB	9/20/2022 15:45	SMPL
N4543-08	YBFC3	1136	0.2472 -		SFAM01.1	PPB	9/20/2022 15:47	SMPL
N4543-09	YBFC5	1192	0.2601 -		SFAM01.1	PPB	9/20/2022 15:49	SMPL
N4543-10	YBFC7	1755	0.3902 -		SFAM01.1	PPB	9/20/2022 15:52	SMPL
N4543-11	YBF95	2118	0.4742 -		SFAM01.1	PPB	9/20/2022 15:54	SMPL
N4543-12	YBFA1	2118	0.4742 -		SFAM01.1	PPB	9/20/2022 15:56	SMPL
N4543-13	YBFA3	2234	0.501 -		SFAM01.1	PPB	9/20/2022 15:58	SMPL
N4543-14	YBFB3	2119	0.4744 -		SFAM01.1	PPB	9/20/2022 16:01	SMPL
PB147774BL	PBW774	145	0.0181 -		SFAM01.1	PPB	9/20/2022 16:06	SMPL
N4544-01	YBF58	449	0.0884 -		SFAM01.1	PPB	9/20/2022 16:08	SMPL
N4544-02	YBF58D	466	0.0923 -		SFAM01.1	PPB	9/20/2022 16:10	SMPL
CCV036	CCV036	20763	4.7839 -		SFAM01.1	PPB	9/20/2022 16:12	SMPL
CCB036	CCB036	51	-0.0036 -		SFAM01.1	PPB	9/20/2022 16:15	SMPL
N4544-03	YBF58S	4830	1.101 -		SFAM01.1	PPB	9/20/2022 16:17	SMPL
N4544-04	YBF60	418	0.0812 -		SFAM01.1	PPB	9/20/2022 16:19	SMPL
N4544-05	YBFB8	411	0.0796 -		SFAM01.1	PPB	9/20/2022 16:21	SMPL
N4544-06	YBFC0	454	0.0895 -		SFAM01.1	PPB	9/20/2022 16:24	SMPL
N4544-07	YBFC2	390	0.0747 -		SFAM01.1	PPB	9/20/2022 16:26	SMPL
N4544-08	YBFC4	423	0.0824 -		SFAM01.1	PPB	9/20/2022 16:28	SMPL
N4544-09	YBFC6	401	0.0773 -		SFAM01.1	PPB	9/20/2022 16:31	SMPL
N4544-10	YBFC8	411	0.0796 -		SFAM01.1	PPB	9/20/2022 16:33	SMPL
N4544-11	YBF96	461	0.0911 -		SFAM01.1	PPB	9/20/2022 16:35	SMPL
N4544-12	YBFA2	411	0.0796 -		SFAM01.1	PPB	9/20/2022 16:37	SMPL
N4544-13	YBFA4	476	0.0946 -		SFAM01.1	PPB	9/20/2022 16:40	SMPL
N4544-14	YBFB4	407	0.0787 -		SFAM01.1	PPB	9/20/2022 16:42	SMPL
CCV037	CCV037	20794	4.7911 -		SFAM01.1	PPB	9/20/2022 16:44	SMPL
CCB037	CCB037	58	-0.002 -		SFAM01.1	PPB	9/20/2022 16:46	SMPL