



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB122096

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	09/21/2022	08:42
2	CAL2	1	Water	NA	NA	20.8	7.00	09/21/2022	08:43
3	CAL3	1	Water	NA	NA	20.8	10.03	09/21/2022	08:44
4	ICV	1	Water	NA	NA	20.7	7.00	09/21/2022	08:45
5	CCV1	1	Water	NA	NA	20.8	2.01	09/21/2022	08:46
6	N4720-01	1	Water	NA	NA	20.9	5.52	09/21/2022	08:49
7	N4720-01DUP	1	Water	NA	NA	20.8	5.53	09/21/2022	08:50
8	N4720-02	1	Water	NA	NA	20.3	5.58	09/21/2022	08:51
9	N4720-03	1	Water	NA	NA	20.4	9.16	09/21/2022	09:00
10	N4726-01	1	Water	NA	NA	21.6	7.90	09/21/2022	09:03
11	N4726-02	1	Water	NA	NA	21.9	9.12	09/21/2022	09:05
12	N4726-03	1	Water	NA	NA	20.7	8.87	09/21/2022	09:10
13	N4726-04	1	Water	NA	NA	21.7	5.99	09/21/2022	09:11
14	N4726-05	1	Water	NA	NA	21.3	6.18	09/21/2022	09:14
15	CCV2	1	Water	NA	NA	20.8	12.03	09/21/2022	09:15

WORKLIST(Hardcopy Internal Chain)

WB 122096

WorkList Name : ph n4672

WorkList ID : 163346

Department : Wet-Chemistry

Date : 09-21-2022 08:35:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/20/2022	Water	N4720-01	pH	D	Cool 4 deg C	PSEG03	N11	ORA-0131	09/20/2022 9040C
09/20/2022	Water	N4720-02	pH	A	Cool 4 deg C	PSEG03	N11	ORA-0094	09/20/2022 9040C
09/20/2022	Water	N4720-03	pH	A	Cool 4 deg C	PSEG03	N11	ORA-0135	09/20/2022 9040C
09/20/2022	Water	N4726-01	pH	A	Cool 4 deg C	PSEG03	N11	CLI-0120	09/20/2022 9040C
09/20/2022	Water	N4726-02	pH	A	Cool 4 deg C	PSEG03	N11	82522	09/20/2022 9040C
09/20/2022	Water	N4726-03	pH	A	Cool 4 deg C	PSEG03	N11	CLI-0144	09/20/2022 9040C
09/20/2022	Water	N4726-04	pH	A	Cool 4 deg C	PSEG03	N11	CLI-0621	09/20/2022 9040C
09/20/2022	Water	N4726-05	pH	C	Cool 4 deg C	PSEG03	N11	CLI-0620	09/20/2022 9040C

Date/Time 09.21.22 08:40

Raw Sample Received by: JD 1629

Raw Sample Relinquished by: CP SR

Date/Time 09.21.22

Raw Sample Received by: CP SR

Raw Sample Relinquished by: JD 1629

10:00

CP SR

JD 1629