



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB122097

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	09/21/2022	08:42
2	CAL2	1	Water	NA	NA	20.8	7.00	09/21/2022	08:43
3	CAL3	1	Water	NA	NA	20.8	10.03	09/21/2022	08:44
4	ICV	1	Water	NA	NA	20.7	7.00	09/21/2022	08:45
5	CCV1	1	Water	NA	NA	20.7	2.01	09/21/2022	09:20
6	N4716-03	1	Water	NA	NA	22.9	7.30	09/21/2022	09:22
7	N4716-03DUP	1	Water	NA	NA	22.8	7.31	09/21/2022	09:23
8	N4716-04	1	Water	NA	NA	22.3	8.23	09/21/2022	09:25
9	CCV2	1	Water	NA	NA	20.8	12.03	09/21/2022	09:26

18122097

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity n4718

WorkList ID : 163347

Department : Wet-Chemistry

Date : 09-21-2022 08:36:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/20/2022	Water	N4716-03	Corrosivity	A	Cool 4 deg C	PSEG03 N11	321	09/20/2022	9040C
09/20/2022	Water	N4716-04	Corrosivity	A	Cool 4 deg C	PSEG03 N11	324	09/20/2022	9040C

Date/Time 09.21.22 08:41 10:10

Raw Sample Received by: 20/02/2022 CP Ser

Raw Sample Relinquished by: 20/02/2022 CP Ser