



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 9/22/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 09/21/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 09/22/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122100

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4723-01	SB-13-7.5-8.0	1	0.91	8.92	9.83	9.46	95.9	
N4723-02	SB-13-COMP	2	0.93	8.59	9.52	9.02	94.2	
N4723-03	SB-14-6.5-7.0	3	0.93	8.77	9.7	8.56	87.0	
N4723-04	SB-14-COMP	4	0.91	9.05	9.96	8.59	84.9	
N4723-05	SB-16-6.5-7.0	5	0.95	9.03	9.98	8.34	81.8	
N4723-06	SB-16-COMP	6	0.91	8.86	9.77	8.09	81.0	
N4723-07	SB-12-6.5-7.0	7	0.91	8.75	9.66	9.18	94.5	
N4723-08	SB-12-COMP	8	0.88	8.86	9.74	8.95	91.1	
N4723-09	SB-15-6.5-7.0	9	0.91	9.04	9.95	8.97	89.2	
N4723-10	SB-15-COMP	10	0.92	9.02	9.94	8.39	82.8	
N4723-11	SB-11-8.5-9.0	11	0.95	9.01	9.96	8.95	88.8	
N4723-12	SB-11-COMP	12	0.92	9.03	9.95	9.55	95.6	
N4723-13	SB-26-8.5-9.0	13	0.91	8.81	9.72	9.14	93.4	
N4723-14	SB-26-COMP	14	0.93	8.73	9.66	9.36	96.6	
N4723-15	SB-27-6.5-7.0	15	0.91	8.75	9.66	8.55	87.3	
N4723-16	SB-27-COMP	16	0.91	8.87	9.78	8.2	82.2	
N4723-17	SB-28-6.5-7.0	17	0.95	9.03	9.98	8.3	81.4	
N4723-18	SB-28-COMP	18	0.91	8.67	9.58	8.32	85.5	
N4723-19	SB-29-6.5-7.0	19	0.88	8.83	9.71	9.02	92.2	
N4723-20	SB-29-COMP	20	0.91	9.06	9.97	8.69	85.9	
N4724-01	SB-30-6.5-7.0	21	0.95	9.00	9.95	8.64	85.4	
N4724-02	SB-30-COMP	22	0.95	8.98	9.93	8.59	85.1	
N4724-03	SB-10-8.5-9.0	23	0.91	8.86	9.77	9.41	95.9	
N4724-04	SB-10-COMP	24	0.95	9.00	9.95	9.04	89.9	
N4724-05	SB-09-11.5-12.0	25	0.93	8.79	9.72	9.09	92.8	
N4724-06	SB-09-COMP	26	0.93	8.59	9.52	9.32	97.7	
N4724-07	SB-01-8.5-9.0	27	0.91	8.94	9.85	8.29	82.6	
N4724-08	SB-01-COMP	28	0.96	9.01	9.97	9.17	91.1	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 9/22/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 09/21/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 09/22/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122100

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4724-09	SB-02-10.5-11.0	29	0.91	8.66	9.57	8.66	89.5	
N4724-10	SB-02-COMP	30	0.96	9.01	9.97	8.65	85.3	
N4724-11	SB-03-11.5-12.0	31	0.95	8.75	9.7	9.09	93.0	
N4724-12	SB-03-COMP	32	0.93	8.89	9.82	8.62	86.5	
N4724-13	SB-04-12.5-13.0	33	0.95	9.03	9.98	9.00	89.1	
N4724-14	SB-04-COMP	34	0.91	8.98	9.89	9.18	92.1	
N4724-15	SB-07-7.5-8.0	35	0.86	8.91	9.77	8.83	89.5	
N4724-16	SB-07-COMP	36	0.92	8.98	9.9	9.22	92.4	
N4731-01	SB-1	43	0.91	8.90	9.81	9.76	99.4	
N4731-02	SB-2	44	0.93	8.70	9.63	9.6	99.7	
N4732-01	WC-1	37	0.92	8.89	9.81	8.66	87.1	
N4732-03	WC-1-EPH-2	38	0.9	8.72	9.62	8.71	89.6	
N4733-01	AU-05-092122	39	0.91	8.84	9.75	9.36	95.6	
N4735-01	40117	45	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4735-02	1823	46	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4735-03	1824	47	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N4738-01	1675	40	0.93	8.90	9.83	9.62	97.6	
N4750-01	SB-15	48	0.93	8.59	9.52	9.4	98.6	
N4750-02	SB-16	49	0.91	9.06	9.97	9.9	99.2	
N4750-03	23417	50	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
N4750-05	34214	51	0.91	8.64	9.55	9.43	98.6	
N4750-07	34216	52	0.93	8.73	9.66	7.77	78.4	
N4752-01	NB-07-092122	41	0.91	9.05	9.96	9.95	99.9	
N4752-02	NB-07-092122-EPH-2	42	0.91	9.05	9.96	9.95	99.9	
N4753-01	EO-01-092122	53	0.91	8.92	9.83	9.21	93.0	
N4753-02	EO-01-092122-EPH-2	54	0.95	9.01	9.96	9.37	93.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$