

LB122103


7471B

INSTRUMENT ID: CV4

Linear

 μ Abs.:

59116



A= 0.0000e+000

B= 1.7113e-004

C= 2.4842e-002

Rho= 0.9995510

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Y/D
0.0	0.000	0.064	0.064	231	0.000	231					-
0.2	0.200	0.236	0.036	1233	0.0 %	1233					18
2.5	2.500	2.416	-0.084	13974	0.0 %	13974					-3
5.0	5.000	5.040	0.040	29309	0.0 %	29309					1
7.5	7.500	7.302	-0.198	42523	0.0 %	42523					-3
10.0	10.000	10.141	0.141	59116	0.0 %	59116					1

LB122103 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Sep 2022 11:12:04

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	231	-	0.0000	7471B	PPB	21 Sep 2022 11:27:19	S	Std
0.2 - 1	502	1233	-	0.2000	7471B	PPB	21 Sep 2022 11:29:36	S	Std
2.5 - 1	52.5	13974	-	2.5000	7471B	PPB	21 Sep 2022 11:31:53	S	Std
5.0 - 1	55	29309	-	5.0000	7471B	PPB	21 Sep 2022 11:36:47	S	Std
7.5 - 1	57.5	42523	-	7.5000	7471B	PPB	21 Sep 2022 11:39:04	S	Std
10.0 - 1	59.5	59116	-	10.0000	7471B	PPB	21 Sep 2022 11:46:30	S	Std
ICV38 - 1	ICV38	23586	4.0611	-	7471B	PPB	21 Sep 2022 11:51:25	U	SMPL
ICB38 - 1	ICB38	-136	0.0016	-	7471B	PPB	21 Sep 2022 11:53:43	U	SMPL
CCV38 - 1	CCV38	30708	5.2799	-	7471B	PPB	21 Sep 2022 11:56:00	U	SMPL
CCB38 - 1	CCB38	-158	-0.0022	-	7471B	PPB	21 Sep 2022 11:58:15	U	SMPL
CRA - 1	CRA	987	0.1937	-	7471B	PPB	21 Sep 2022 12:00:32	U	SMPL
HighStd - 1	HighStd	56891	9.7606	-	7471B	PPB	21 Sep 2022 12:02:48	U	SMPL
ChkStd - 1	ChkStd	39862	6.8464	-	7471B	PPB	21 Sep 2022 12:05:03	U	SMPL
PB147783BL - 1	PBS	127	0.0466	-	7471B	PPB	21 Sep 2022 12:07:22	U	SMPL
PB147783BS - 1	LCSS	21573	3.7166	-	7471B	PPB	21 Sep 2022 12:09:40	U	SMPL
N4695-02 - 1	INCINERATOR-ASH	6118	1.0718	-	7471B	PPB	21 Sep 2022 12:11:56	U	SMPL
N4700-01 - 1	ROLLOFF-SOIL	2904	0.5218	-	7471B	PPB	21 Sep 2022 12:14:15	U	SMPL
N4701-01 - 1	OK-02-091622	2154	0.3935	-	7471B	PPB	21 Sep 2022 12:16:32	U	SMPL
N4702-01 - 1	BU-03-091622	3242	0.5796	-	7471B	PPB	21 Sep 2022 12:18:49	U	SMPL
N4703-02 - 1	ETGI-363	3630	0.6460	-	7471B	PPB	21 Sep 2022 12:21:04	U	SMPL
CCV39 - 1	CCV39	30647	5.2694	-	7471B	PPB	21 Sep 2022 12:23:19	U	SMPL
CCB39 - 1	CCB39	-187	-0.0072	-	7471B	PPB	21 Sep 2022 12:25:34	U	SMPL
N4712-01 - 1	HR-01-091922	12988	2.2475	-	7471B	PPB	21 Sep 2022 12:27:51	U	SMPL
N4713-01 - 1	OR-03-091922	4096	0.7258	-	7471B	PPB	21 Sep 2022 12:30:07	U	SMPL
N4714-01 - 1	OK-01-091922	2878	0.5174	-	7471B	PPB	21 Sep 2022 12:32:23	U	SMPL
N4714-01DUP - 1	OK-01-091922DUP	2625	0.4741	-	7471B	PPB	21 Sep 2022 12:36:46	U	SMPL
N4714-01MS - 1	OK-01-091922MS	24746	4.2596	-	7471B	PPB	21 Sep 2022 12:39:01	U	SMPL
N4714-01MSD - 1	OK-01-091922MSD	22461	3.8686	-	7471B	PPB	21 Sep 2022 12:41:18	U	SMPL
N4714-01LX5 - 1		435	0.0993	-	7471B	PPB	21 Sep 2022 12:43:36	U	SMPL
N4714-01A - 1		24824	4.2730	-	7471B	PPB	21 Sep 2022 12:45:54	U	SMPL
CCV40 - 1	CCV40	31572	5.4277	-	7471B	PPB	21 Sep 2022 12:48:11	U	SMPL
CCB40 - 1	CCB40	78	0.0382	-	7471B	PPB	21 Sep 2022 12:50:29	U	SMPL