



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB122112

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	09/22/2022	08:20
2	CAL2	1	Water	NA	NA	20.8	7.00	09/22/2022	08:21
3	CAL3	1	Water	NA	NA	20.8	10.03	09/22/2022	08:22
4	ICV	1	Water	NA	NA	20.8	7.00	09/22/2022	08:25
5	CCV1	1	Water	NA	NA	20.8	2.01	09/22/2022	08:30
6	N4750-09	1	Water	NA	NA	22.6	4.95	09/22/2022	08:31
7	N4750-09DUP	1	Water	NA	NA	22.7	4.96	09/22/2022	08:32
8	N4750-11	1	Water	NA	NA	23.9	7.22	09/22/2022	08:35
9	CCV2	1	Water	NA	NA	20.7	12.03	09/22/2022	08:36

WORKLIST(Hardcopy Internal Chain)

122112

WorkList Name : ph n4750

WorkList ID : 163395

Department : Wet-Chemistry

Date : 09-22-2022 08:15:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/21/2022	Water	N4750-09	pH	Cool 4 deg C	PSEG03	N11	23418	09/21/2022	9040C
09/21/2022	Water	N4750-11	pH	Cool 4 deg C	PSEG03	N11	34215	09/21/2022	9040C

Date/Time 09/22/22 08:15
 Raw Sample Received by: 7D(WC)
 Raw Sample Relinquished by: CP SW

Date/Time 09/22/22 10:30
 Raw Sample Received by: CP SW
 Raw Sample Relinquished by: 7A(WC)