



TOTAL SUSPENDED SOLIDS - SM2540D

TEMP1 IN: 103 °C 09/22/2022 13:00 TEMP1 OUT: 103 °C 09/22/2022 14:00
TEMP2 IN: 103 °C 09/22/2022 14:30 TEMP2 OUT: 104 °C 09/22/2022 15:30
TEMP3 IN: 104 °C 09/23/2022 09:00 TEMP3 OUT: 103 °C 09/23/2022 10:30
TEMP4 IN: 103 °C 09/23/2022 11:00 TEMP4 OUT: 103 °C 09/23/2022 12:30

SUPERVISOR: Iwona

ANALYST: jignesh

Date: 09/22/2022

Run Number: LB122118

BalanceID: WC SC-6

OvenID: WC OVEN-1

FilterID: 17416528

ThermometerID: WET OVEN#1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB122118BL	LB122118BL	1.3596	1.3596	100	1.3597	1.3597	1.3597	0.0001	1
2	LB122118BS	LB122118BS	1.4857	1.4857	100	1.5387	1.5387	1.5387	0.0530	530
3	N4727-02	COMP	1.4102	1.4102	50	1.4284	1.4284	1.4284	0.0182	364
4	N4742-07	TWP-05	1.4191	1.4191	100	1.5070	1.5070	1.5070	0.0879	879
5	N4742-08	TWP-14	1.4260	1.4260	250	1.4747	1.4747	1.4747	0.0487	194.8
6	N4742-08DUP	TWP-14DUP	1.4195	1.4195	250	1.4683	1.4683	1.4683	0.0488	195.2
7	N4742-09	TWP-21	1.4073	1.4073	100	1.6103	1.6103	1.6103	0.2030	2030
8	N4756-01	PIT	1.4217	1.4217	200	1.4367	1.4367	1.4367	0.0150	75
9	N4758-01	PIT	1.4237	1.4237	500	1.4497	1.4497	1.4497	0.0260	52
10	N4759-01	PIT	1.4106	1.4106	300	1.4338	1.4338	1.4338	0.0232	77.3

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

13122218

WorkList Name : tss n4756

WorkList ID : 163433

Department : Wet-Chemistry

Date : 09-23-2022 08:02:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/28/2022	Water	N4727-02	TSS	Cool 4 deg C	ARAM01	N11	COMP	09/21/2022	SM2540 D
09/28/2022	Water	N4742-07	TSS	Cool 4 deg C	LIRO01	N11	TWP-05	09/21/2022	SM2540 D
09/28/2022	Water	N4742-08	TSS	Cool 4 deg C	LIRO01	N11	TWP-14	09/21/2022	SM2540 D
09/28/2022	Water	N4742-09	TSS	Cool 4 deg C	LIRO01	N11	TWP-21	09/21/2022	SM2540 D
09/29/2022	Water	N4756-01	TSS	Cool 4 deg C	URBA02	N11	PIT	09/22/2022	SM2540 D
09/29/2022	Water	N4758-01	TSS	Cool 4 deg C	URBA02	N11	PIT	09/22/2022	SM2540 D
09/29/2022	Water	N4759-01	TSS	Cool 4 deg C	URBA02	N11	PIT	09/22/2022	SM2540 D

Date/Time 09/23/22 08:10
 Raw Sample Received by: JDL
 Raw Sample Relinquished by: CP SR

Date/Time 09/23/22 11:00
 Raw Sample Received by: CP SR
 Raw Sample Relinquished by: JDL