


LB122123

SFAM01.1 INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.029	0.029	88	0.000	88					11
0.05	0.050					185	0				-6
0.20	0.200	0.221	0.021	810	0.0 %	810					3
2.50	2.500	2.343	-0.157	8788	0.0 %	8788					0
5.00	5.000	5.128	0.128	19260	0.0 %	19260					-1
7.50	7.500	7.537	0.037	28319	0.0 %	28319					
10.0	10.000	9.943	-0.057	37365	0.0 %	37365					

LB122123
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0	88	-		0 SFAM01.1	PPB	9/22/2022 14:12	Std
	0.2	810	-		0.2 SFAM01.1	PPB	9/22/2022 14:14	Std
	2.5	8788	-		2.5 SFAM01.1	PPB	9/22/2022 14:19	Std
	5	19260	-		5 SFAM01.1	PPB	9/22/2022 14:23	Std
	7.5	28319	-		7.5 SFAM01.1	PPB	9/22/2022 14:28	Std
	10	37365	-		10 SFAM01.1	PPB	9/22/2022 14:32	Std
ICV040	ICV040	14891	3.9658 -		SFAM01.1	PPB	9/22/2022 14:38	SMPL
ICB040	ICB040	17	0.0101 -		SFAM01.1	PPB	9/22/2022 14:40	SMPL
CCV043	CCV043	19894	5.2963 -		SFAM01.1	PPB	9/22/2022 14:42	SMPL
CCB043	CCB043	1	0.0059 -		SFAM01.1	PPB	9/22/2022 14:45	SMPL
PB147844BL	PBW844	28	0.013 -		SFAM01.1	PPB	9/22/2022 14:47	SMPL
N4606-01	YBF50	-8	0.0035 -		SFAM01.1	PPB	9/22/2022 14:49	SMPL
N4606-02	YBF50D	-54	-0.0088 -		SFAM01.1	PPB	9/22/2022 14:51	SMPL
N4606-03	YBF50S	4452	1.1896 -		SFAM01.1	PPB	9/22/2022 14:54	SMPL
N4606-04	YBF52	-75	-0.0144 -		SFAM01.1	PPB	9/22/2022 14:56	SMPL
N4606-05	YBF54	7383	1.9691 -		SFAM01.1	PPB	9/22/2022 14:58	SMPL
N4606-06	YBF55	7267	1.9382 -		SFAM01.1	PPB	9/22/2022 15:00	SMPL
N4606-07	YBF56	14010	3.7315 -		SFAM01.1	PPB	9/22/2022 15:03	SMPL
N4606-08	YBFE1	-152	-0.0348 -		SFAM01.1	PPB	9/22/2022 15:05	SMPL
N4606-09	MYBFC9	-33	-0.0032 -		SFAM01.1	PPB	9/22/2022 15:07	SMPL
N4606-10	MYBFD1	-50	-0.0077 -		SFAM01.1	PPB	9/22/2022 15:10	SMPL
N4606-11	MYBFD3	-78	-0.0152 -		SFAM01.1	PPB	9/22/2022 15:12	SMPL
N4606-12	MYBFD5	-65	-0.0117 -		SFAM01.1	PPB	9/22/2022 15:14	SMPL
N4606-13	MYBFD7	-66	-0.012 -		SFAM01.1	PPB	9/22/2022 15:16	SMPL
N4606-14	YBF93	-32	-0.0029 -		SFAM01.1	PPB	9/22/2022 15:19	SMPL
N4606-15	YBF97	-64	-0.0114 -		SFAM01.1	PPB	9/22/2022 15:21	SMPL
N4606-16	YBF99	-92	-0.0189 -		SFAM01.1	PPB	9/22/2022 15:23	SMPL
N4606-17	YBFA5	-85	-0.017 -		SFAM01.1	PPB	9/22/2022 15:25	SMPL
N4606-18	MYBFD9	-73	-0.0138 -		SFAM01.1	PPB	9/22/2022 15:28	SMPL
N4606-19	YBF69	-64	-0.0114 -		SFAM01.1	PPB	9/22/2022 15:30	SMPL
CCV044	CCV044	19861	5.2875 -		SFAM01.1	PPB	9/22/2022 15:32	SMPL
CCB044	CCB044	24	0.012 -		SFAM01.1	PPB	9/22/2022 15:34	SMPL
N4606-20	YBFA7	-68	-0.0125 -		SFAM01.1	PPB	9/22/2022 15:37	SMPL
PB147845BL	PBW845	21	0.0112 -		SFAM01.1	PPB	9/22/2022 15:39	SMPL
N4607-01	YBF51	-68	-0.0125 -		SFAM01.1	PPB	9/22/2022 15:41	SMPL
N4607-02	YBF51D	-82	-0.0162 -		SFAM01.1	PPB	9/22/2022 15:44	SMPL
N4607-03	YBF51S	4398	1.1752 -		SFAM01.1	PPB	9/22/2022 15:46	SMPL
N4607-04	YBF53	-113	-0.0245 -		SFAM01.1	PPB	9/22/2022 15:48	SMPL
N4607-05	MYBFD0	-79	-0.0154 -		SFAM01.1	PPB	9/22/2022 15:50	SMPL
N4607-06	MYBFD2	-52	-0.0082 -		SFAM01.1	PPB	9/22/2022 15:53	SMPL
N4607-07	MYBFD4	-72	-0.0136 -		SFAM01.1	PPB	9/22/2022 15:55	SMPL
N4607-08	MYBFD6	-59	-0.0101 -		SFAM01.1	PPB	9/22/2022 15:57	SMPL
N4607-09	MYBFD8	37	0.0154 -		SFAM01.1	PPB	9/22/2022 15:59	SMPL
N4607-10	YBF94	105	0.0335 -		SFAM01.1	PPB	9/22/2022 16:02	SMPL
N4607-11	YBF98	47	0.0181 -		SFAM01.1	PPB	9/22/2022 16:04	SMPL
N4607-12	YBFA0	10	0.0082 -		SFAM01.1	PPB	9/22/2022 16:06	SMPL

LB122123

INSTRUMENT ID : CV1

N4607-13	YBFA6	-9	0.0032 -	SFAM01.1	PPB	9/22/2022 16:08 SMPL
N4607-14	MYBFEO	43	0.017 -	SFAM01.1	PPB	9/22/2022 16:11 SMPL
N4607-15	YBF70	36	0.0152 -	SFAM01.1	PPB	9/22/2022 16:13 SMPL
N4607-16	YBFA8	26	0.0125 -	SFAM01.1	PPB	9/22/2022 16:15 SMPL
PB147846BL	PBW846	80	0.0269 -	SFAM01.1	PPB	9/22/2022 16:18 SMPL
N4015-02	MDL-WATER	216	0.063 -	SFAM01.1	PPB	9/22/2022 16:25 SMPL
CCV045	CCV045	20262	5.3942 -	SFAM01.1	PPB	9/22/2022 16:27 SMPL
CCB045	CCB045	-173	-0.0404 -	SFAM01.1	PPB	9/22/2022 16:29 SMPL