



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: jignesh

Run Number: LB122141

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP99050
calibration std. phosphate 0.5 ppm	WP99049
calibration std. phosphate 0.3 ppm	WP99048
calibration std. phosphate 0.1 ppm	WP99047
calibration std. phosphate 0.05 ppm	WP99046
calibration std. 0 ppm	WP99045
phosphate CCV std.	WP99051
Combined reagent	WP99054
Phenolphthalein indicator	WP96861
Sodium hydroxide, 1N	WP98644
Phosphate ICV-LCS Std	WP99052

Intercept: -0.0008

Slope: 0.6506

Regression: 0.999974

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000			09/23/2022	12:40
2	CAL2	0.05	1	50	50	0.033			09/23/2022	12:40
3	CAL3	0.10	1	50	50	0.065			09/23/2022	12:41
4	CAL4	0.30	1	50	50	0.191			09/23/2022	12:41
5	CAL5	0.50	1	50	50	0.324			09/23/2022	12:42
6	CAL6	1.00	1	50	50	0.651			09/23/2022	12:43



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: jignesh

Run Number: LB122141

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.321	0.495	09/23/2022	12:43
2	ICB		1	50	50	0.001	0.003	09/23/2022	12:44
3	CCV1	0.50	1	50	50	0.324	0.499	09/23/2022	12:44
4	CCB1		1	50	50	0.001	0.003	09/23/2022	12:45
5	PB147843BL		1	50	50	0.002	0.004	09/23/2022	12:45
6	PB147843BS	0.50	1	50	50	0.318	0.490	09/23/2022	12:46
7	N4635-01		1	50	50	1.048	1.612	09/23/2022	12:46
8	N4636-01		1	50	50	0.561	0.864	09/23/2022	12:47
9	N4696-01		1	50	50	2.048	3.149	09/23/2022	12:47
10	N4756-01		1	50	50	1.750	2.691	09/23/2022	12:48
11	N4756-01DUP		1	50	50	1.762	2.709	09/23/2022	12:48
12	N4756-01MS	0.50	1	50	50	1.924	2.958	09/23/2022	12:49
13	N4756-01MSD	0.50	1	50	50	1.958	3.011	09/23/2022	12:49
14	N4758-01		1	50	50	1.685	2.591	09/23/2022	12:50
15	CCV2	0.50	1	50	50	0.323	0.498	09/23/2022	12:50
16	CCB2		1	50	50	0.001	0.003	09/23/2022	12:51
17	N4635-01		5	50	50	0.564	0.868	09/23/2022	12:51
18	N4696-01		20	50	50	0.553	0.851	09/23/2022	12:52
19	N4756-01		100	50	50	0.478	0.736	09/23/2022	12:52
20	N4756-01DUP		100	50	50	0.478	0.736	09/23/2022	12:53
21	N4756-01MS	0.50	100	50	50	0.480	0.739	09/23/2022	12:53
22	N4756-01MSD	0.50	100	50	50	0.480	0.739	09/23/2022	12:54
23	N4758-01		20	50	50	0.368	0.567	09/23/2022	12:54
24	CCV3	0.50	1	50	50	0.322	0.496	09/23/2022	12:55
25	CCB3		1	50	50	0.002	0.004	09/23/2022	12:55