

245.1


60122146

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44196



A= 0.0000e+000

B= 2.2710e-004

C= 1.2008e-002

Rho= 0.9998729

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	W.D
0.0	0.000	0.017	0.017	21	0.000	21					6
0.05	0.050	0.053	0.003	181	0.0 %	181	0				1
0.2	0.200	0.202	0.002	837	0.0 %	837					2
2.5	2.500	2.454	-0.046	10751	0.0 %	10751					2
5.0	5.000	5.086	0.086	22342	0.0 %	22342					1
7.5	7.500	7.389	-0.111	32484	0.0 %	32484					0
10.0	10.000	10.049	0.049	44196	0.0 %	44196					

LB122146 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 23 Sep 2022 14:21:35

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	21	-	0.0000	245.1	PPB	23 Sep 2022 14:34:44	S	Std
0.05 - 1	50.05	181	-	0.0500	245.1	PPB	23 Sep 2022 14:37:01	S	Std
0.2 - 1	50.2	837	-	0.2000	245.1	PPB	23 Sep 2022 14:39:18	S	Std
2.5 - 1	52.5	10751	-	2.5000	245.1	PPB	23 Sep 2022 14:41:35	S	Std
5.0 - 1	55	22342	-	5.0000	245.1	PPB	23 Sep 2022 14:43:51	S	Std
7.5 - 1	57.5	32484	-	7.5000	245.1	PPB	23 Sep 2022 14:53:57	S	Std
10.0 - 1	510	44196	-	10.0000	245.1	PPB	23 Sep 2022 15:03:16	S	Std
ICV42 - 1	ICV42	18127	4.1287	-245.1		PPB	23 Sep 2022 15:06:03	U	SMPL
ICB42 - 1	ICB42	-32	0.0047	-245.1		PPB	23 Sep 2022 15:08:18	U	SMPL
CCV53 - 1	CCV53	22479	5.1171	-245.1		PPB	23 Sep 2022 15:10:34	U	SMPL
CCB53 - 1	CCB53	-21	0.0072	-245.1		PPB	23 Sep 2022 15:12:50	U	SMPL
CRA - 1	CRA	172	0.0511	-245.1		PPB	23 Sep 2022 15:15:07	U	SMPL
HighStd - 1	HighStd	44533	10.1257	-245.1		PPB	23 Sep 2022 15:17:21	U	SMPL
ChkStd - 1	ChkStd	30828	7.0132	-245.1		PPB	23 Sep 2022 15:19:37	U	SMPL
PB147877BL - 1	PBW	39	0.0209	-245.1		PPB	23 Sep 2022 15:24:54	U	SMPL
PB147877BS - 1	LCSW	18156	4.1353	-245.1		PPB	23 Sep 2022 15:27:10	U	SMPL
N4666-02 - 1	FALL-2022	2783	0.6440	-245.1		PPB	23 Sep 2022 15:29:26	U	SMPL
N4671-02 - 1	EFF-WW	1115	0.2652	-245.1		PPB	23 Sep 2022 15:31:45	U	SMPL
N4742-07 - 1	TWP-05	2495	0.5786	-245.1		PPB	23 Sep 2022 15:34:02	U	SMPL
N4742-08 - 1	TWP-14	2138	0.4976	-245.1		PPB	23 Sep 2022 15:36:18	U	SMPL
N4742-09 - 1	TWP-21	2419	0.5614	-245.1		PPB	23 Sep 2022 15:38:34	U	SMPL
CCV54 - 1	CCV54	22234	5.0615	-245.1		PPB	23 Sep 2022 15:40:49	U	SMPL
CCB54 - 1	CCB54	-24	0.0066	-245.1		PPB	23 Sep 2022 15:43:05	U	SMPL
N4766-01 - 1	EFFLUENT-DAY-1-METALS	340	0.0892	-245.1		PPB	23 Sep 2022 15:45:21	U	SMPL
N4766-01DUP - 1	EFFLUENT-DAY-1-METALSDUP	314	0.0833	-245.1		PPB	23 Sep 2022 15:47:36	U	SMPL
N4766-01MS - 1	EFFLUENT-DAY-1-METALSMS	15751	3.5891	-245.1		PPB	23 Sep 2022 15:49:52	U	SMPL
N4766-01MSD - 1	EFFLUENT-DAY-1-METALSMSD	16477	3.7540	-245.1		PPB	23 Sep 2022 15:55:06	U	SMPL
N4766-01LX5 - 1		96	0.0338	-245.1		PPB	23 Sep 2022 15:57:21	U	SMPL
N4766-01A - 1		17519	3.9907	-245.1		PPB	23 Sep 2022 15:59:38	U	SMPL
CCV55 - 1	CCV55	22109	5.0331	-245.1		PPB	23 Sep 2022 16:04:09	U	SMPL
CCB55 - 1	CCB55	-51	0.0004	-245.1		PPB	23 Sep 2022 16:06:25	U	SMPL