



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 9/26/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:30
In Date: 09/24/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:25
Out Date: 09/25/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122158

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4741-01	EY533	1	0.92	8.72	9.64	7.79	78.8	
N4741-02	EY534	2	0.92	8.99	9.91	8.03	79.1	
N4741-03	EY535	3	0.92	9.02	9.94	8.1	79.6	
N4741-04	EY536	4	0.91	8.72	9.63	8.21	83.7	
N4741-05	EY537	5	0.91	9.01	9.92	8.06	79.4	
N4741-06	EY538	6	0.92	8.79	9.71	7.74	77.6	
N4741-07	EY539	7	0.91	8.79	9.7	7.51	75.1	
N4741-08	EY540	8	0.9	8.99	9.89	8.11	80.2	
N4741-09	EY541	9	0.91	8.97	9.88	8.28	82.2	
N4741-10	EY542	10	0.92	9.00	9.92	8.08	79.6	
N4741-11	EY543	11	0.9	9.09	9.99	7.73	75.1	
N4741-12	EY544	12	0.9	8.79	9.69	7.7	77.4	
N4741-13	EY544D	13	0.9	8.79	9.69	7.7	77.4	
N4741-14	EY544S	14	0.9	8.79	9.69	7.7	77.4	
N4741-15	EY545	15	0.9	9.08	9.98	7.97	77.9	
N4741-16	EY546	16	0.9	9.09	9.99	8.04	78.5	
N4741-17	EY547	17	0.9	9.02	9.92	8.02	78.9	
N4741-18	EY548	18	0.9	8.71	9.61	7.6	76.9	
N4741-19	EY549	19	0.91	8.94	9.85	8.03	79.6	
N4741-20	EY550	20	0.9	8.87	9.77	8.2	82.3	
N4741-21	EY551	21	0.92	8.86	9.78	8.41	84.5	
N4741-22	EY552	22	0.9	9.09	9.99	7.82	76.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$