



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 9/26/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 13:40
In Date: 09/24/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:30
Out Date: 09/25/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122159

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4743-01	EY553	1	0.9	8.92	9.82	7.76	76.9	
N4743-02	EY554	2	0.9	8.97	9.87	8.02	79.4	
N4743-03	EY555	3	0.92	8.97	9.89	8.17	80.8	
N4743-04	EY556	4	0.91	8.88	9.79	8.17	81.8	
N4743-05	EY557	5	0.9	8.85	9.75	8.18	82.3	
N4743-06	EY558	6	0.9	9.09	9.99	8.49	83.5	
N4743-07	EY559	7	0.9	8.88	9.78	8.14	81.5	
N4743-08	EY560	8	0.9	9.09	9.99	8.43	82.8	
N4743-09	EY561	9	0.9	9.07	9.97	7.82	76.3	
N4743-10	EY562	10	0.9	9.04	9.94	7.87	77.1	
N4743-11	EY563	11	0.9	8.73	9.63	7.75	78.5	
N4743-12	EY564	12	0.91	8.95	9.86	8.02	79.4	
N4743-13	EY565	13	0.9	8.90	9.8	8.01	79.9	
N4743-14	EY566	14	0.91	9.07	9.98	8.02	78.4	
N4743-15	EY567	15	0.91	8.72	9.63	7.72	78.1	
N4743-16	EY567D	16	0.91	8.72	9.63	7.72	78.1	
N4743-17	EY567S	17	0.91	8.72	9.63	7.72	78.1	
N4743-18	EY568	18	0.92	9.07	9.99	7.91	77.1	
N4743-19	EY569	19	0.91	9.01	9.92	7.89	77.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$