



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 9/26/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 15:15
In Date: 09/24/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:48
Out Date: 09/25/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122163

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4773-01	EY5B1	1	0.91	8.84	9.75	8.11	81.4	
N4773-02	EY5B2	2	0.91	8.70	9.61	7.51	75.9	
N4773-03	EY5B3	3	0.91	8.76	9.67	8.05	81.5	
N4773-04	EY5B4	4	0.91	8.97	9.88	8.26	81.9	
N4773-05	EY5B6	5	0.9	8.88	9.78	7.89	78.7	
N4773-06	EY5B7	6	0.91	9.00	9.91	8.08	79.7	
N4773-07	EY5B8	7	0.9	8.83	9.73	7.42	73.8	
N4773-08	EY5B9	8	0.91	8.75	9.66	7.57	76.1	
N4773-09	EY5C0	9	0.9	9.07	9.97	7.58	73.6	
N4773-10	EY5C1	10	0.91	8.83	9.74	7.77	77.7	
N4773-11	EY5C2	11	0.92	9.05	9.97	7.49	72.6	
N4773-12	EY5C3	12	0.91	8.82	9.73	8.09	81.4	
N4773-13	EY5C4	13	0.9	8.95	9.85	8.41	83.9	
N4773-14	EY5C5	14	0.91	9.08	9.99	8.5	83.6	
N4773-15	EY5C6	15	0.93	8.81	9.74	8.27	83.3	
N4773-16	EY5C7	16	0.91	8.79	9.7	8.38	85.0	
N4773-17	EY5C8	17	0.92	8.79	9.71	8.34	84.4	
N4773-18	EY5C9	18	0.92	8.97	9.89	8.46	84.1	
N4773-19	EY5D0	19	0.91	8.88	9.79	8.09	80.9	
N4773-20	EY5D8	20	0.9	9.02	9.92	8.82	87.8	
N4773-21	EY5D8D	21	0.9	9.02	9.92	8.82	87.8	
N4773-22	EY5D8S	22	0.9	9.02	9.92	8.82	87.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$