



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 9/26/2022

OVENTEMP IN Celsius(°C): 108
Time IN: 15:35
In Date: 09/24/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:55
Out Date: 09/25/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122165

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4775-01	EY5H0	1	0.9	8.89	9.79	8.24	82.6	
N4775-02	EY5H1	2	0.92	8.82	9.74	8.16	82.1	
N4775-03	EY5H2	3	0.93	9.06	9.99	8.29	81.2	
N4775-04	EY5H3	4	0.92	8.75	9.67	8.02	81.1	
N4775-05	EY5H4	5	0.92	8.83	9.75	8.03	80.5	
N4775-06	EY5H5	6	0.92	8.85	9.77	7.88	78.6	
N4775-07	EY5H6	7	0.92	9.05	9.97	8.14	79.8	
N4775-08	EY5H7	8	0.92	9.07	9.99	7.89	76.8	
N4775-09	EY5H8	9	0.91	9.03	9.94	7.95	78.0	
N4775-10	EY5H9	10	0.92	8.93	9.85	8.1	80.4	
N4775-11	EY5J0	11	0.91	8.87	9.78	7.87	78.5	
N4775-12	EY5J1	12	0.93	8.71	9.64	7.66	77.3	
N4775-13	EY5J2	13	0.92	8.86	9.78	8.00	79.9	
N4775-14	EY5J3	14	0.91	8.98	9.89	7.89	77.7	
N4775-15	EY5J4	15	0.91	8.84	9.75	7.84	78.4	
N4775-16	EY5J4D	16	0.91	8.84	9.75	7.84	78.4	
N4775-17	EY5J4S	17	0.91	8.84	9.75	7.84	78.4	
N4775-18	EY5J5	18	0.92	9.02	9.94	7.54	73.4	
N4775-19	EY5J6	19	0.92	9.07	9.99	7.58	73.4	
N4775-20	EY5J7	20	0.92	9.02	9.94	7.54	73.4	
N4775-21	EY5J8	21	0.92	9.04	9.96	8.28	81.4	
N4775-22	EY5J9	22	0.92	8.93	9.85	8.16	81.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$