



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 9/26/2022

OVENTEMP IN Celsius(°C): 108
Time IN: 16:20
In Date: 09/24/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:12
Out Date: 09/25/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122170

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4797-01	EY5R2	1	0.92	8.85	9.77	7.95	79.4	
N4797-02	EY5R3	2	0.91	8.89	9.8	8.07	80.5	
N4797-03	EY5R4	3	0.9	9.09	9.99	8.09	79.1	
N4797-04	EY5R5	4	0.91	9.03	9.94	8.06	79.2	
N4797-05	EY5R7	5	0.92	9.05	9.97	8.1	79.3	
N4797-06	EY5R8	6	0.91	9.02	9.93	7.7	75.3	
N4797-08	EY5S0	7	0.9	9.07	9.97	8.12	79.6	
N4797-09	EY5S1	8	0.92	9.00	9.92	7.87	77.2	
N4797-10	EY5S2	9	0.91	9.07	9.98	8.05	78.7	
N4797-11	EY5S3	10	0.92	8.86	9.78	7.7	76.5	
N4797-12	EY5S4	11	0.91	9.04	9.95	8.08	79.3	
N4797-13	EY5S6	12	0.91	8.94	9.85	8.02	79.5	
N4797-14	EY5S6	13	0.9	8.94	9.84	7.99	79.3	
N4797-15	EY5S7	14	0.92	8.88	9.8	7.87	78.3	
N4797-16	EY5S8	15	0.91	8.82	9.73	7.99	80.3	
N4797-17	EY5S9	16	0.9	8.87	9.77	7.99	79.9	
N4797-18	EY5T0	17	0.91	8.78	9.69	7.92	79.8	
N4797-19	EY5T1	18	0.9	9.00	9.9	8.26	81.8	
N4797-20	EY5T3	19	0.91	9.04	9.95	8.32	82.0	
N4797-21	EY5T3D	20	0.91	9.04	9.95	8.32	82.0	
N4797-22	EY5T3S	21	0.91	9.04	9.95	8.32	82.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$