



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 9/26/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:15
In Date: 09/25/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 09/26/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122174

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4799-01	EY5X3	1	0.91	8.86	9.77	8.17	81.9	
N4799-02	EY5X4	2	0.91	8.75	9.66	8.26	84.0	
N4799-03	EY5X5	3	0.91	8.68	9.59	8.03	82.0	
N4799-04	EY5X6	4	0.9	8.84	9.74	8.17	82.2	
N4799-05	EY5X7	5	0.93	8.67	9.6	8.03	81.9	
N4799-06	EY5X8	6	0.91	8.74	9.65	8.14	82.7	
N4799-07	EY5X9	7	0.91	8.91	9.82	8.47	84.8	
N4799-08	EY5Y0	8	0.93	8.82	9.75	8.25	83.0	
N4799-09	EY5Y1	9	0.9	8.72	9.62	7.97	81.1	
N4799-10	EY5Y2	10	0.91	9.04	9.95	8.62	85.3	
N4799-11	EY5Y2D	11	0.91	9.04	9.95	8.62	85.3	
N4799-12	EY5Y2S	12	0.91	9.04	9.95	8.62	85.3	
N4799-13	EY5Y3	13	0.93	9.02	9.95	8.83	87.6	
N4799-14	EY5Y4	14	0.95	8.99	9.94	8.99	89.4	
N4799-15	EY5Y5	15	0.91	8.89	9.8	8.64	87.0	
N4799-16	EY5Y6	16	0.93	9.03	9.96	8.52	84.1	
N4799-17	EY5Y7	17	0.93	8.93	9.86	8.33	82.9	
N4799-18	EY5Y8	18	0.89	8.72	9.61	8.12	82.9	
N4799-19	EY5Y9	19	0.93	8.91	9.84	8.45	84.4	
N4799-20	EY5Z0	20	0.88	8.64	9.52	8.05	83.0	
N4799-21	EY5Z1	21	0.93	9.01	9.94	8.14	80.0	
N4799-22	EY5Z2	22	0.88	8.78	9.66	7.94	80.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$