



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 9/27/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:05
In Date: 09/26/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:08
Out Date: 09/27/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122178

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4801-01	EY612	1	0.91	9.07	9.98	8.16	79.9	
N4801-02	EY613	2	0.93	8.98	9.91	7.84	76.9	
N4801-03	EY614	3	0.95	9.01	9.96	8.28	81.4	
N4801-04	EY615	4	0.91	9.06	9.97	8.07	79.0	
N4801-05	EY616	5	0.92	8.92	9.84	8.07	80.2	
N4801-06	EY617	6	0.91	9.05	9.96	8.64	85.4	
N4801-07	EY618	7	0.92	8.63	9.55	8.45	87.3	
N4801-08	EY619	8	0.91	8.64	9.55	7.8	79.7	
N4801-09	EY620	9	0.91	8.71	9.62	8.03	81.7	
N4801-10	EY621	10	0.92	8.82	9.74	8.11	81.5	
N4801-11	EY622	11	0.91	8.70	9.61	8.33	85.3	
N4801-12	EY623	12	0.89	8.75	9.64	8.13	82.7	
N4801-13	EY624	13	0.91	8.69	9.6	7.81	79.4	
N4801-14	EY625	14	0.93	8.68	9.61	7.93	80.6	
N4801-15	EY626	15	0.91	8.70	9.61	8.00	81.5	
N4801-16	EY627	16	0.94	8.67	9.61	8.16	83.3	
N4801-17	EY628	17	0.95	9.01	9.96	8.49	83.7	
N4801-18	EY629	18	0.91	8.76	9.67	8.5	86.6	
N4801-19	EY631	19	0.96	9.02	9.98	9.16	90.9	
N4801-20	EY632	20	0.91	8.91	9.82	8.96	90.3	
N4801-21	EY632D	21	0.91	8.91	9.82	8.96	90.3	
N4801-22	EY632S	22	0.91	8.91	9.82	8.96	90.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$