



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 9/27/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:15
In Date: 09/26/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 09/27/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122179

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4802-01	EY633	1	0.95	9.01	9.96	8.71	86.1	
N4802-02	EY634	2	0.93	9.04	9.97	8.24	80.9	
N4802-03	EY635	3	0.94	9.03	9.97	8.85	87.6	
N4802-04	EY636	4	0.91	8.90	9.81	8.24	82.4	
N4802-05	EY637	5	0.95	8.91	9.86	8.54	85.2	
N4802-06	EY638	6	0.91	8.81	9.72	7.75	77.6	
N4802-07	EY639	7	0.91	8.97	9.88	7.95	78.5	
N4802-08	EY640	8	0.91	8.91	9.82	7.94	78.9	
N4802-09	EY641	9	0.95	9.02	9.97	8.49	83.6	
N4802-10	EY642	10	0.95	8.59	9.54	7.93	81.3	
N4802-11	EY643	11	0.96	9.01	9.97	8.44	83.0	
N4802-12	EY644	12	0.89	8.99	9.88	7.92	78.2	
N4802-13	EY645	13	0.86	9.01	9.87	7.87	77.8	
N4802-14	EY646	14	0.93	8.94	9.87	8.03	79.4	
N4802-15	EY647	15	0.91	8.64	9.55	8.2	84.4	
N4802-16	EY648	16	0.91	8.94	9.85	8.18	81.3	
N4802-17	EY649	17	0.93	8.72	9.65	8.15	82.8	
N4802-18	EY650	18	0.91	8.71	9.62	8.06	82.1	
N4802-19	EY651	19	0.91	9.05	9.96	8.5	83.9	
N4802-20	EY653	20	0.95	8.91	9.86	8.21	81.5	
N4802-21	EY653D	21	0.95	8.91	9.86	8.21	81.5	
N4802-22	EY653S	22	0.95	8.91	9.86	8.21	81.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$