



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Jignesh

Parameter: Flash Point

Supervisor Review By: Sohil

Run Number: LB122181

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	83.0	09/23/2022	09:15
2	N4720-06		1	14	100.00	>212.0	>212.0	09/23/2022	09:45
3	N4750-03		1	15	100.00	>212.0	>212.0	09/23/2022	10:15
4	N4750-03DUP		1	13	100.00	>212.0	>212.0	09/23/2022	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : flash point-oil-0923

WorkList ID : 163470

Department : Wet-Chemistry

Date : 09-23-2022 08:13:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/23/2022	Solid	N4720-06	Flash Point	Cool 4 deg C	PSEG03	N11	5922	09/20/2022	1010B
09/28/2022	Solid	N4750-03	Flash Point	Cool 4 deg C	PSEG03	N11	34217	09/21/2022	1010B

Date/Time 9/23/2022 8:15
 Raw Sample Received by: SJ (wc)
 Raw Sample Relinquished by: CP (SM)

Date/Time 9/23/2022 12:00
 Raw Sample Received by: CP (SM)
 Raw Sample Relinquished by: SJ (wc)