

7471B


LB122183

INSTRUMENT ID: 011

Linear

 μ Abs.:

54239



A= 0.0000e+000

B= 1.8355e-004

C= 2.0770e-002

Rho= 0.9999715

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/100
0.0	0.000	0.021	0.021	2	0.000	2					
0.2	0.200	0.182	-0.018	880	0.0 %	880					19
2.5	2.500	2.463	-0.037	13304	0.0 %	13304					11
5.0	5.000	5.041	0.041	27351	0.0 %	27351					10
7.5	7.500	7.516	0.016	40836	0.0 %	40836					0
10.0	10.000	9.976	-0.024	54239	0.0 %	54239					0

LB122183 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Sep 2022 11:19:59

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SP	2	-	0.0000	7471B	PPB	26 Sep 2022 11:32:35	S	Std
0.2 - 1	SP-2	880	-	0.2000	7471B	PPB	26 Sep 2022 11:34:52	S	Std
2.5 - 1	SP-5	13304	-	2.5000	7471B	PPB	26 Sep 2022 11:37:09	S	Std
5.0 - 1	SP-8	27351	-	5.0000	7471B	PPB	26 Sep 2022 11:39:31	S	Std
7.5 - 1	SP-10	40836	-	7.5000	7471B	PPB	26 Sep 2022 11:41:50	S	Std
10.0 - 1	SP-12	54239	-	10.0000	7471B	PPB	26 Sep 2022 11:44:07	S	Std
ICV43 - 1	ICV43	20584	3.7990	-7471B	PPB	26 Sep 2022 11:46:54	U	SMPL	
ICB43 - 1	ICB43	-118	-0.0009	-7471B	PPB	26 Sep 2022 11:49:09	U	SMPL	
CCV56 - 1	CCV56	26383	4.8634	-7471B	PPB	26 Sep 2022 11:51:26	U	SMPL	
CCB56 - 1	CCB56	-145	-0.0058	-7471B	PPB	26 Sep 2022 11:53:42	U	SMPL	
CRA - 1	CRA	1076	0.2183	-7471B	PPB	26 Sep 2022 11:55:59	U	SMPL	
HighStd - 1	HighStd	53283	9.8010	-7471B	PPB	26 Sep 2022 11:58:14	U	SMPL	
ChkStd - 1	ChkStd	38637	7.1127	-7471B	PPB	26 Sep 2022 12:00:29	U	SMPL	
PB147887BL - 1	PBS	-136	-0.0042	-7471B	PPB	26 Sep 2022 12:02:48	U	SMPL	
PB147887BS - 1	LCSS	19344	3.5714	-7471B	PPB	26 Sep 2022 12:05:06	U	SMPL	
N4732-01 - 1	WC-1	6798	1.2686	-7471B	PPB	26 Sep 2022 12:07:22	U	SMPL	
N4733-01 - 1	AU-05-092122	2397	0.4607	-7471B	PPB	26 Sep 2022 12:09:40	U	SMPL	
N4738-01 - 1	1675	363	0.0874	-7471B	PPB	26 Sep 2022 12:11:57	U	SMPL	
N4752-01 - 1	NB-07-092122	1093	0.2214	-7471B	PPB	26 Sep 2022 12:14:14	U	SMPL	
N4753-01 - 1	EO-01-092122	6498	1.2135	-7471B	PPB	26 Sep 2022 12:16:30	U	SMPL	
CCV57 - 1	CCV57	27381	5.0466	-7471B	PPB	26 Sep 2022 12:18:45	U	SMPL	
CCB57 - 1	CCB57	-128	-0.0027	-7471B	PPB	26 Sep 2022 12:21:00	U	SMPL	
N4764-02 - 1	354	1048	0.2131	-7471B	PPB	26 Sep 2022 12:23:17	U	SMPL	
N4764-02DUP - 1	354DUP	934	0.1922	-7471B	PPB	26 Sep 2022 12:25:33	U	SMPL	
N4764-02MS - 1	354MS	23681	4.3675	-7471B	PPB	26 Sep 2022 12:27:48	U	SMPL	
N4764-02MSD - 1	354MSD	22948	4.2329	-7471B	PPB	26 Sep 2022 12:30:04	U	SMPL	
N4764-05 - 1	RBR-01	1119	0.2262	-7471B	PPB	26 Sep 2022 12:32:22	U	SMPL	
N4764-08 - 1	355	1613	0.3168	-7471B	PPB	26 Sep 2022 12:34:40	U	SMPL	
N4764-11 - 1	R2	4421	0.8323	-7471B	PPB	26 Sep 2022 12:36:57	U	SMPL	
N4765-01 - 1	OK-01-092222	1767	0.3451	-7471B	PPB	26 Sep 2022 12:39:14	U	SMPL	
N4780-01 - 1	SP-1	1210	0.2429	-7471B	PPB	26 Sep 2022 12:41:31	U	SMPL	
N4780-11 - 1	SP-2	1071	0.2174	-7471B	PPB	26 Sep 2022 12:43:47	U	SMPL	
CCV58 - 1	CCV58	27430	5.0556	-7471B	PPB	26 Sep 2022 12:46:03	U	SMPL	
CCB58 - 1	CCB58	-74	0.0072	-7471B	PPB	26 Sep 2022 12:48:18	U	SMPL	
N4780-21 - 1	SP-3	1891	0.3679	-7471B	PPB	26 Sep 2022 12:50:36	U	SMPL	
N4780-31 - 1	SP-4	1209	0.2427	-7471B	PPB	26 Sep 2022 12:52:51	U	SMPL	
N4782-01 - 1	SASP2-T-1-(0-6)	58817	10.8167	-7471B	PPB	26 Sep 2022 12:55:06	U	SMPL	
N4789-01 - 1	RB20515	696	0.1485	-7471B	PPB	26 Sep 2022 12:57:22	U	SMPL	
N4809-01 - 1	TP1-0923	38104	7.0148	-7471B	PPB	26 Sep 2022 12:59:41	U	SMPL	
N4809-04 - 1	TP2-0923	104939	19.2825	-7471B	PPB	26 Sep 2022 13:01:58	U	SMPL	
N4809-07 - 1	TP3-0923	24070	4.4389	-7471B	PPB	26 Sep 2022 13:04:17	U	SMPL	
N4764-02LX5 - 1		52	0.0303	-7471B	PPB	26 Sep 2022 13:06:39	U	SMPL	
N4764-02A - 1		23199	4.2790	-7471B	PPB	26 Sep 2022 13:08:58	U	SMPL	
N4782-01DLX2 - 1		29036	5.3504	-7471B	PPB	26 Sep 2022 13:11:14	U	SMPL	
CCV59 - 1	CCV59	27762	5.1165	-7471B	PPB	26 Sep 2022 13:13:33	U	SMPL	
CCB59 - 1	CCB59	-115	-0.0003	-7471B	PPB	26 Sep 2022 13:15:50	U	SMPL	
N4809-04DLX5 - 1		24641	4.5437	-7471B	PPB	26 Sep 2022 13:18:07	U	SMPL	
CCV60 - 1	CCV60	27268	5.0259	-7471B	PPB	26 Sep 2022 13:20:23	U	SMPL	
CCB60 - 1	CCB60	-108	0.0009	-7471B	PPB	26 Sep 2022 13:22:40	U	SMPL	