

LB122189


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4622E



A= 0.0000e+000

B= 2.1747e-004 slope

C= -9.2212e-003
→ m/accept

Rho= 0.9993125

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0, 1)
0.00	0.000	0.007	0.007	74	0.000	74					11
0.05	0.050					185	0				-12
0.20	0.200	0.176	-0.024	852	0.0 %	852					-2
2.50	2.500	2.447	-0.053	11293	0.0 %	11293					-2
5.00	5.000	5.244	0.244	24158	0.0 %	24158					5
7.50	7.500	7.283	-0.217	33531	0.0 %	33531					-3
10.0	10.000	10.043	0.043	46226	0.0 %	46226					0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	74	-		0 SFAM01.1	PPB	9/26/2022 15:28	Std
	0.2 S01	852	-		0.2 SFAM01.1	PPB	9/26/2022 15:30	Std
	2.5 S02	11293	-		2.5 SFAM01.1	PPB	9/26/2022 15:32	Std
	5 S03	24158	-		5 SFAM01.1	PPB	9/26/2022 15:49	Std
	7.5 S04	33531	-		7.5 SFAM01.1	PPB	9/26/2022 15:58	Std
	10 S05	46226	-		10 SFAM01.1	PPB	9/26/2022 16:01	Std
ICV044	ICV044	18132	3.9339 -		SFAM01.1	PPB	9/26/2022 16:03	SMPL
ICB044	ICB044	-2	-0.0097 -		SFAM01.1	PPB	9/26/2022 16:06	SMPL
CCV061	CCV061	23807	5.168 -		SFAM01.1	PPB	9/26/2022 16:08	SMPL
CCB061	CCB061	-63	-0.0229 -		SFAM01.1	PPB	9/26/2022 16:10	SMPL
PB147908BL	PBW908	57	0.0032 -		SFAM01.1	PPB	9/26/2022 16:12	SMPL
N4624-01	MCB8F7	905	0.1876 -		SFAM01.1	PPB	9/26/2022 16:15	SMPL
N4624-02	MCB8G5	33	-0.002 -		SFAM01.1	PPB	9/26/2022 16:17	SMPL
N4624-03	MCB8G7	684	0.1395 -		SFAM01.1	PPB	9/26/2022 16:19	SMPL
N4624-04	MCB8G9	22	-0.0044 -		SFAM01.1	PPB	9/26/2022 16:21	SMPL
N4624-05	MCB8H2	23	-0.0042 -		SFAM01.1	PPB	9/26/2022 16:24	SMPL
N4624-06	MCB8H4	106	0.0138 -		SFAM01.1	PPB	9/26/2022 16:26	SMPL
N4624-07	MCB8H8	125	0.018 -		SFAM01.1	PPB	9/26/2022 16:28	SMPL
N4624-08	MCB8H9	32	-0.0023 -		SFAM01.1	PPB	9/26/2022 16:31	SMPL
N4624-09	MCB8J0	6	-0.0079 -		SFAM01.1	PPB	9/26/2022 16:33	SMPL
N4624-12	MCB8J8	-16	-0.0127 -		SFAM01.1	PPB	9/26/2022 16:35	SMPL
N4624-13	MCB8J9	64	0.0047 -		SFAM01.1	PPB	9/26/2022 16:37	SMPL
N4624-14	MCB8K3	44	0.0003 -		SFAM01.1	PPB	9/26/2022 16:40	SMPL
PB147910BL	PBW910	83	0.0088 -		SFAM01.1	PPB	9/26/2022 16:42	SMPL
N4625-01	MCB8F8	19	-0.0051 -		SFAM01.1	PPB	9/26/2022 16:44	SMPL
N4625-02	MCB8G6	96	0.0117 -		SFAM01.1	PPB	9/26/2022 16:46	SMPL
N4625-03	MCB8G8	58	0.0034 -		SFAM01.1	PPB	9/26/2022 16:49	SMPL
N4625-04	MCB8H1	20	-0.0049 -		SFAM01.1	PPB	9/26/2022 16:51	SMPL
N4625-05	MCB8H5	20	-0.0049 -		SFAM01.1	PPB	9/26/2022 16:53	SMPL
N4625-06	MCB8H6	-57	-0.0216 -		SFAM01.1	PPB	9/26/2022 16:55	SMPL
CCV062	CCV062	23350	5.0686 -		SFAM01.1	PPB	9/26/2022 16:58	SMPL
CCB062	CCB062	32	-0.0023 -		SFAM01.1	PPB	9/26/2022 17:00	SMPL
N4625-07	MCB8J1	119	0.0167 -		SFAM01.1	PPB	9/26/2022 17:02	SMPL
N4625-08	MCB8J2	108	0.0143 -		SFAM01.1	PPB	9/26/2022 17:04	SMPL
N4625-09	MCB8J3	-24	-0.0144 -		SFAM01.1	PPB	9/26/2022 17:07	SMPL
N4625-10	MCB8K1	-79	-0.0264 -		SFAM01.1	PPB	9/26/2022 17:09	SMPL
N4625-11	MCB8K2	-8	-0.011 -		SFAM01.1	PPB	9/26/2022 17:11	SMPL
N4625-12	MCB8K4	-44	-0.0188 -		SFAM01.1	PPB	9/26/2022 17:14	SMPL
PB147911BL	PBW911	42	-0.0001 -		SFAM01.1	PPB	9/26/2022 17:16	SMPL
N4519-01	MDL-WATER	245	0.0441 -		SFAM01.1	PPB	9/26/2022 17:18	SMPL
CCV063	CCV063	23477	5.0963 -		SFAM01.1	PPB	9/26/2022 17:23	SMPL
CCB063	CCB063	32	-0.0023 -		SFAM01.1	PPB	9/26/2022 17:25	SMPL