



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB122194

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	09/26/2022	16:22
2	CAL2	1	Water	NA	NA	20.7	7.00	09/26/2022	16:23
3	CAL3	1	Water	NA	NA	20.7	10.03	09/26/2022	16:24
4	ICV	1	Water	NA	NA	20.7	6.98	09/26/2022	16:25
5	CCV1	1	Water	NA	NA	20.7	2.01	09/26/2022	16:26
6	N4763-01	1	Water	NA	NA	20.2	7.23	09/26/2022	16:30
7	N4792-01	1	Water	NA	NA	20.1	4.82	09/26/2022	16:32
8	N4821-01	1	Water	NA	NA	22.8	6.74	09/26/2022	16:35
9	N4827-01	1	Water	NA	NA	22.1	6.83	09/26/2022	16:38
10	N4833-01	1	Water	NA	NA	20.6	7.77	09/26/2022	16:40
11	N4833-02	1	Water	NA	NA	20.8	8.16	09/26/2022	16:42
12	N4835-01	1	Water	NA	NA	20.7	8.43	09/26/2022	16:45
13	N4835-02	1	Water	NA	NA	20.2	8.40	09/26/2022	16:48
14	N4837-01	1	Water	NA	NA	20.6	8.31	09/26/2022	16:50
15	N4837-02	1	Water	NA	NA	20.8	8.35	09/26/2022	16:52
16	CCV2	1	Water	NA	NA	20.8	12.03	09/26/2022	17:00
17	N4839-01	1	Water	NA	NA	21.1	7.92	09/26/2022	17:02
18	N4839-02	1	Water	NA	NA	21.0	7.89	09/26/2022	17:05
19	N4841-01	1	Water	NA	NA	21.2	8.12	09/26/2022	17:10
20	N4841-02	1	Water	NA	NA	21.3	8.13	09/26/2022	17:12

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	N4843-01	1	Water	NA	NA	22.6	6.64	09/26/2022	17:15
22	N4843-01DUP	1	Water	NA	NA	22.7	6.65	09/26/2022	17:16
23	CCV3	1	Water	NA	NA	20.7	2.02	09/26/2022	17:20

VB 122194

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N4827

WorkList ID : 163519

Department : Wet-Chemistry

Date : 09-26-2022 14:37:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/21/2022	Water	N4763-01	pH	Cool 4 deg C	PSEG03	WC-GW		09/21/2022	9040C
09/20/2022	Water	N4792-01	pH	Cool 4 deg C	STER15	N11	POOL-WATER	09/20/2022	9040C
09/26/2022	Water	N4821-01	pH	Cool 4 deg C	PSEG03	N11	NBS-SUMP-WATER-0926	09/26/2022	9040C
09/26/2022	Water	N4827-01	pH	Cool 4 deg C	PSEG03	N11	OVS-OUTLET	09/26/2022	9040C
09/22/2022	Water	N4833-01	pH	Cool 4 deg C	ENVI70	M11	004-1	09/22/2022	9040C
09/22/2022	Water	N4833-02	pH	Cool 4 deg C	ENVI70	M11	004-2	09/22/2022	9040C
09/22/2022	Water	N4835-01	pH	Cool 4 deg C	ENVI70	N11	529317-3	09/22/2022	9040C
09/22/2022	Water	N4835-02	pH	Cool 4 deg C	ENVI70	N11	529317-4	09/22/2022	9040C
09/22/2022	Water	N4837-01	pH	Cool 4 deg C	ENVI70	N11	251563-3	09/22/2022	9040C
09/22/2022	Water	N4837-02	pH	Cool 4 deg C	ENVI70	N11	251563-4	09/22/2022	9040C
09/22/2022	Water	N4839-01	pH	Cool 4 deg C	ENVI70	N11	5592-1	09/22/2022	9040C
09/22/2022	Water	N4839-02	pH	Cool 4 deg C	ENVI70	N11	5592-2	09/22/2022	9040C
09/23/2022	Water	N4841-01	pH	Cool 4 deg C	ENVI70	N11	252672-5	09/23/2022	9040C
09/23/2022	Water	N4841-02	pH	Cool 4 deg C	ENVI70	N11	252672-6	09/23/2022	9040C
09/26/2022	Water	N4843-01	pH	Cool 4 deg C	PSEG03	N11	HD-SUMPWATER-0926	09/26/2022	9040C

Date/Time 09/26/22 16:20
 Raw Sample Received by: 7P(WC)
 Raw Sample Relinquished by: RJ CFFA-1646

Date/Time 09/26/22 18:00
 Raw Sample Received by: RJ CFFA-1646
 Raw Sample Relinquished by: 7P(WC)