

LB122196 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Sep 2022 08:40:41

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	65	-	0.0000	7470A	PPB	27 Sep 2022 08:47:47	S	Std
0.2 - 1	51.2	827	-	0.2000	7470A	PPB	27 Sep 2022 08:50:04	S	Std
2.5 - 1	52.5	11619	-	2.5000	7470A	PPB	27 Sep 2022 08:52:22	S	Std
5.0 - 1	55	21018	-	5.0000	7470A	PPB	27 Sep 2022 08:54:39	S	Std
7.5 - 1	57.1	31151	-	7.5000	7470A	PPB	27 Sep 2022 08:56:55	S	Std
10.0 - 1	510	43422	-	10.0000	7470A	PPB	27 Sep 2022 08:59:12	S	Std
ICV45 - 1	ICV45	16470	3.8375	-	7470A	PPB	27 Sep 2022 09:04:14	U	SMPL
ICB45 - 1	ICB45	-125	-0.0512	-	7470A	PPB	27 Sep 2022 09:06:29	U	SMPL
CCV64 - 1	CCV64	21087	4.9194	-	7470A	PPB	27 Sep 2022 09:08:45	U	SMPL
CCB64 - 1	CCB64	-108	-0.0472	-	7470A	PPB	27 Sep 2022 09:11:01	U	SMPL
CRA - 1	CRA	1037	0.2211	-	7470A	PPB	27 Sep 2022 09:15:51	U	SMPL
HighStd - 1	HighStd	43246	10.1119	-	7470A	PPB	27 Sep 2022 09:18:07	U	SMPL
ChkStd - 1	ChkStd	29649	6.9257	-	7470A	PPB	27 Sep 2022 09:20:23	U	SMPL
PB147905BL - 1	PBW	-144	-0.0557	-	7470A	PPB	27 Sep 2022 09:22:40	U	SMPL
PB147905BS - 1	LCSW	17014	3.9650	-	7470A	PPB	27 Sep 2022 09:24:58	U	SMPL
N4724-02 - 1	SB-30-COMP	-55	-0.0348	-	7470A	PPB	27 Sep 2022 09:27:14	U	SMPL
N4724-04 - 1	SB-10-COMP	13	-0.0189	-	7470A	PPB	27 Sep 2022 09:29:31	U	SMPL
N4724-06 - 1	SB-09-COMP	17	-0.0179	-	7470A	PPB	27 Sep 2022 09:31:48	U	SMPL
N4724-08 - 1	SB-01-COMP	3	-0.0212	-	7470A	PPB	27 Sep 2022 09:34:05	U	SMPL
N4724-10 - 1	SB-02-COMP	22	-0.0168	-	7470A	PPB	27 Sep 2022 09:36:20	U	SMPL
CCV65 - 1	CCV65	22029	5.1401	-	7470A	PPB	27 Sep 2022 09:38:36	U	SMPL
CCB65 - 1	CCB65	-80	-0.0407	-	7470A	PPB	27 Sep 2022 09:40:51	U	SMPL
N4724-12 - 1	SB-03-COMP	6	-0.0205	-	7470A	PPB	27 Sep 2022 09:43:08	U	SMPL
N4724-14 - 1	SB-04-COMP	29	-0.0151	-	7470A	PPB	27 Sep 2022 09:45:23	U	SMPL
N4724-16 - 1	SB-07-COMP	24	-0.0163	-	7470A	PPB	27 Sep 2022 09:47:39	U	SMPL
N4742-02 - 1	SB-08-COMP	31	-0.0147	-	7470A	PPB	27 Sep 2022 09:49:55	U	SMPL
N4742-04 - 1	SB-06-COMP	6	-0.0205	-	7470A	PPB	27 Sep 2022 09:52:11	U	SMPL
N4742-06 - 1	SB-05-COMP	22	-0.0168	-	7470A	PPB	27 Sep 2022 09:54:28	U	SMPL
N4750-07 - 1	34216	39	-0.0128	-	7470A	PPB	27 Sep 2022 09:56:44	U	SMPL
N4764-03 - 1	354	11	-0.0193	-	7470A	PPB	27 Sep 2022 09:59:01	U	SMPL
N4764-06 - 1	RBR-01	30	-0.0149	-	7470A	PPB	27 Sep 2022 10:01:18	U	SMPL
N4764-09 - 1	355	34	-0.0140	-	7470A	PPB	27 Sep 2022 10:03:34	U	SMPL
CCV66 - 1	CCV66	21913	5.1129	-	7470A	PPB	27 Sep 2022 10:05:49	U	SMPL
CCB66 - 1	CCB66	-83	-0.0414	-	7470A	PPB	27 Sep 2022 10:08:05	U	SMPL
N4764-12 - 1	R2	6	-0.0205	-	7470A	PPB	27 Sep 2022 10:10:21	U	SMPL
N4780-02 - 1	SP-1	20	-0.0172	-	7470A	PPB	27 Sep 2022 10:12:37	U	SMPL
N4780-12 - 1	SP-2	17	-0.0179	-	7470A	PPB	27 Sep 2022 10:14:52	U	SMPL
N4780-12DUP - 1	SP-2DUP	38	-0.0130	-	7470A	PPB	27 Sep 2022 10:17:08	U	SMPL
N4780-12MS - 1	SP-2MS	15465	3.6020	-	7470A	PPB	27 Sep 2022 10:19:24	U	SMPL
N4780-12MSD - 1	SP-2MSD	15261	3.5542	-	7470A	PPB	27 Sep 2022 10:21:41	U	SMPL
PB147906BL - 1	PBW	0	-0.0219	-	7470A	PPB	27 Sep 2022 10:23:58	U	SMPL
PB147906BS - 1	LCSW	15701	3.6573	-	7470A	PPB	27 Sep 2022 10:31:16	U	SMPL
N4780-22 - 1	SP-3	-42	-0.0318	-	7470A	PPB	27 Sep 2022 10:33:33	U	SMPL
N4780-32 - 1	SP-4	23	-0.0165	-	7470A	PPB	27 Sep 2022 10:35:49	U	SMPL
CCV67 - 1	CCV67	21513	5.0192	-	7470A	PPB	27 Sep 2022 10:38:04	U	SMPL
CCB67 - 1	CCB67	-67	-0.0376	-	7470A	PPB	27 Sep 2022 10:40:20	U	SMPL
N4790-02 - 1	250FERRYBLVD-SOIL-1	65	-0.0067	-	7470A	PPB	27 Sep 2022 10:42:36	U	SMPL
N4809-02 - 1	TP1-0923	8	-0.0200	-	7470A	PPB	27 Sep 2022 10:44:52	U	SMPL
N4809-05 - 1	TP2-0923	90	-0.0008	-	7470A	PPB	27 Sep 2022 10:47:07	U	SMPL
N4809-08 - 1	TP3-0923	31	-0.0147	-	7470A	PPB	27 Sep 2022 10:49:23	U	SMPL
N4809-08DUP - 1	TP3-0923DUP	58	-0.0083	-	7470A	PPB	27 Sep 2022 10:51:39	U	SMPL
N4809-08MS - 1	TP3-0923MS	15500	3.6102	-	7470A	PPB	27 Sep 2022 10:53:55	U	SMPL
N4809-08MSD - 1	TP3-0923MSD	15891	3.7018	-	7470A	PPB	27 Sep 2022 10:56:12	U	SMPL
PB147841TB - 1	PB147841TB	-73	-0.0390	-	7470A	PPB	27 Sep 2022 10:58:28	U	SMPL
PB147874TB - 1	PB147874TB	45	-0.0114	-	7470A	PPB	27 Sep 2022 11:00:46	U	SMPL
CCV68 - 1	CCV68	21421	4.9977	-	7470A	PPB	27 Sep 2022 11:03:02	U	SMPL
CCB68 - 1	CCB68	-94	-0.0440	-	7470A	PPB	27 Sep 2022 11:05:18	U	SMPL
N4780-12LX5 - 1		-79	-0.0404	-	7470A	PPB	27 Sep 2022 11:07:35	U	SMPL
N4780-12A - 1		17052	3.9739	-	7470A	PPB	27 Sep 2022 11:09:51	U	SMPL
N4809-08LX5 - 1		-156	-0.0585	-	7470A	PPB	27 Sep 2022 11:12:06	U	SMPL
N4809-08A - 1		17111	3.9877	-	7470A	PPB	27 Sep 2022 11:14:23	U	SMPL
CCV69 - 1	CCV69	21072	4.9159	-	7470A	PPB	27 Sep 2022 11:31:26	U	SMPL
CCB69 - 1	CCB69	-79	-0.0404	-	7470A	PPB	27 Sep 2022 11:33:43	U	SMPL