



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB122197

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	10	28.00	82.4	82.7	09/27/2022	09:13
2	N4763-01		1	13	100.00	>212.0	>212.0	09/27/2022	09:35
3	N4821-01		1	14	100.00	>212.0	>212.0	09/27/2022	10:05
4	N4821-01DUP		1	13	100.00	>212.0	>212.0	09/27/2022	10:35
5	N4843-01		1	15	100.00	>212.0	>212.0	09/27/2022	11:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB/22/197

WORKLIST(Hardcopy Internal Chain)

WorkList Name : flash point-0927

WorkList ID : 163541

Department : Wet-Chemistry

Date : 09-27-2022 08:53:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/22/2022	Water	N4763-01	Flash Point	Cool 4 deg C	PSEG03	WC-GW		09/21/2022	1010B
10/03/2022	Water	N4821-01	Flash Point	Cool 4 deg C	PSEG03	N11	NBS-SUMP-WATER-0926	09/26/2022	1010B
10/03/2022	Water	N4843-01	Flash Point	Cool 4 deg C	PSEG03	N11	HD-SUMPWATER-0926	09/26/2022	1010B
10/03/2022	Water	N4943-04	Flash Point	Cool 4 deg C	PSEG03	N11	HD-SUMPWATER-0926	09/26/2022	1010B

SJ

9/27/22

Date/Time 9/27/22 9:00
Raw Sample Received by: SJ(WC)
Raw Sample Relinquished by: CP(SM)

Date/Time 9/27/22 12:00
Raw Sample Received by: CP(SM)
Raw Sample Relinquished by: SJ(WC)