



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB122222

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	09/27/2022	13:42
2	CAL2	1	Water	NA	NA	20.8	7.00	09/27/2022	13:43
3	CAL3	1	Water	NA	NA	20.7	10.03	09/27/2022	13:45
4	ICV	1	Water	NA	NA	20.7	6.98	09/27/2022	13:46
5	CCV1	1	Water	NA	NA	20.8	2.01	09/27/2022	13:48
6	N4812-02	1	Solid	20.02	20	21.4	10.64	09/27/2022	13:50
7	N4812-02DUP	1	Solid	20.03	20	21.5	10.65	09/27/2022	13:51
8	N4850-02	1	Solid	20.04	20	22.4	8.13	09/27/2022	14:00
9	CCV2	1	Water	NA	NA	20.6	12.03	09/27/2022	14:05

WORKLIST(Hardcopy Internal Chain)

09/12/22

WorkList Name : CORROSIONITY N4850 WorkList ID : 163572 Department : Wet-Chemistry Date : 09-27-2022 13:34:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/23/2022	Solid	N4812-02	Corrosivity	Cool 4 deg C	PSEG03	N11	WC-3	09/23/2022	9045D
09/26/2022	Solid	N4850-02	Corrosivity	Cool 4 deg C	ROUX02	N11	WC-14B-SP-2	09/26/2022	9045D

Date/Time 09/12/22 13:40
 Raw Sample Received by: DCC
 Raw Sample Relinquished by: GJ Sm

Date/Time 09/12/22 17:00
 Raw Sample Received by: CP Sm
 Raw Sample Relinquished by: JP/CC