

LB122229


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4542E



A= 0.0000e+000

B= 2.2026e-004

C= 5.2544e-002

Rho= 0.9990497

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01
0.0	0.000	0.065	0.065	55	0.000	55					
0.2	0.200	0.218	0.018	751	0.0 %	751					
2.5	2.500	2.307	-0.193	10236	0.0 %	10236					
5.0	5.000	5.259	0.259	23640	0.0 %	23640					
7.5	7.500	7.293	-0.207	32873	0.0 %	32873					
10.0	10.000	10.058	0.058	45426	0.0 %	45426					

LB122229 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Sep 2022 09:11:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	55	-	0.0000	7470A	PPB	28 Sep 2022 09:55:13	S	Std
0.2 - 1	50.2	751	-	0.2000	7470A	PPB	28 Sep 2022 09:57:30	S	Std
2.5 - 1	52.5	10236	-	2.5000	7470A	PPB	28 Sep 2022 09:59:47	S	Std
5.0 - 1	55	23640	-	5.0000	7470A	PPB	28 Sep 2022 10:02:04	S	Std
7.5 - 1	57.5	32873	-	7.5000	7470A	PPB	28 Sep 2022 10:04:20	S	Std
10.0 - 1	510	45426	-	10.0000	7470A	PPB	28 Sep 2022 10:06:37	S	Std
ICV47 - 1	ICV47	16592	3.7070	-7470A	PPB	28 Sep 2022 10:09:45	U	SMPL	
ICB47 - 1	ICB47	-58	0.0398	-7470A	PPB	28 Sep 2022 10:12:00	U	SMPL	
CCV73 - 1	CCV73	20613	4.5927	-7470A	PPB	28 Sep 2022 10:14:16	U	SMPL	
CCB73 - 1	CCB73	-67	0.0378	-7470A	PPB	28 Sep 2022 10:16:31	U	SMPL	
CRA - 1	CRA	712	0.2094	-7470A	PPB	28 Sep 2022 10:18:48	U	SMPL	
HighStd - 1	HighStd	44361	9.8233	-7470A	PPB	28 Sep 2022 10:21:03	U	SMPL	
ChkStd - 1	ChkStd	30374	6.7426	-7470A	PPB	28 Sep 2022 10:23:19	U	SMPL	
PB147953BL - 1	PBW	-153	0.0188	-7470A	PPB	28 Sep 2022 10:25:37	U	SMPL	
PB147953BS - 1	LCSW	15812	3.5352	-7470A	PPB	28 Sep 2022 10:27:55	U	SMPL	
N4812-02 - 1	WC-3	-57	0.0400	-7470A	PPB	28 Sep 2022 10:30:11	U	SMPL	
N4812-02DUP - 1	WC-3DUP	40	0.0614	-7470A	PPB	28 Sep 2022 10:32:28	U	SMPL	
N4812-02MS - 1	WC-3MS	17560	3.9202	-7470A	PPB	28 Sep 2022 10:34:45	U	SMPL	
N4812-02MSD - 1	WC-3MSD	16080	3.5943	-7470A	PPB	28 Sep 2022 10:37:02	U	SMPL	
PB147954BL - 1	PBW	-81	0.0347	-7470A	PPB	28 Sep 2022 10:39:17	U	SMPL	
CCV74 - 1	CCV74	21762	4.8458	-7470A	PPB	28 Sep 2022 10:41:34	U	SMPL	
CCB74 - 1	CCB74	-87	0.0334	-7470A	PPB	28 Sep 2022 10:43:49	U	SMPL	
PB147954BS - 1	LCSW	17439	3.8936	-7470A	PPB	28 Sep 2022 10:46:06	U	SMPL	
N4821-03 - 1	NBS-SUMP-WATER-0926	-74	0.0362	-7470A	PPB	28 Sep 2022 10:48:22	U	SMPL	
N4821-03DUP - 1	NBS-SUMP-WATER-0926DUP	58	0.0653	-7470A	PPB	28 Sep 2022 10:50:39	U	SMPL	
N4821-03MS - 1	NBS-SUMP-WATER-0926MS	16625	3.7143	-7470A	PPB	28 Sep 2022 10:52:55	U	SMPL	
N4821-03MSD - 1	NBS-SUMP-WATER-0926MSD	15851	3.5438	-7470A	PPB	28 Sep 2022 10:55:10	U	SMPL	
N4843-04 - 1	HD-SUMPWATER-0926	-81	0.0347	-7470A	PPB	28 Sep 2022 10:57:27	U	SMPL	
PB147955BL - 1	PBW	7	0.0541	-7470A	PPB	28 Sep 2022 10:59:45	U	SMPL	
PB147955BS - 1	LCSW	16608	3.7106	-7470A	PPB	28 Sep 2022 11:02:02	U	SMPL	
N4763-01 - 1	WC-GW	222	0.1014	-7470A	PPB	28 Sep 2022 11:04:22	U	SMPL	
CCV75 - 1	CCV75	21825	4.8596	-7470A	PPB	28 Sep 2022 11:08:54	U	SMPL	
CCB75 - 1	CCB75	-140	0.0217	-7470A	PPB	28 Sep 2022 11:11:12	U	SMPL	
N4821-01 - 1	NBS-SUMP-WATER-0926	246	0.1067	-7470A	PPB	28 Sep 2022 11:18:26	U	SMPL	
N4821-01DUP - 1	NBS-SUMP-WATER-0926DUP	182	0.0926	-7470A	PPB	28 Sep 2022 11:20:41	U	SMPL	
N4821-01MS - 1	NBS-SUMP-WATER-0926MS	14644	3.2780	-7470A	PPB	28 Sep 2022 11:22:57	U	SMPL	
N4821-01MSD - 1	NBS-SUMP-WATER-0926MSD	15053	3.3681	-7470A	PPB	28 Sep 2022 11:25:12	U	SMPL	
N4833-01 - 1	004-1	-172	0.0147	-7470A	PPB	28 Sep 2022 11:27:28	U	SMPL	
N4833-02 - 1	004-2	-86	0.0336	-7470A	PPB	28 Sep 2022 11:29:45	U	SMPL	
N4835-01 - 1	529317-3	-97	0.0312	-7470A	PPB	28 Sep 2022 11:32:02	U	SMPL	
N4835-02 - 1	529317-4	-100	0.0305	-7470A	PPB	28 Sep 2022 11:34:19	U	SMPL	
N4837-01 - 1	251563-3	-78	0.0354	-7470A	PPB	28 Sep 2022 11:36:36	U	SMPL	
N4837-02 - 1	251563-4	-87	0.0334	-7470A	PPB	28 Sep 2022 11:38:52	U	SMPL	
CCV76 - 1	CCV76	21110	4.7021	-7470A	PPB	28 Sep 2022 11:41:08	U	SMPL	
CCB76 - 1	CCB76	-94	0.0318	-7470A	PPB	28 Sep 2022 11:43:23	U	SMPL	
N4839-01 - 1	5592-1	-97	0.0312	-7470A	PPB	28 Sep 2022 11:45:40	U	SMPL	
N4839-02 - 1	5592-2	-95	0.0316	-7470A	PPB	28 Sep 2022 11:47:55	U	SMPL	
N4841-01 - 1	252672-5	-62	0.0389	-7470A	PPB	28 Sep 2022 11:50:11	U	SMPL	
N4841-02 - 1	252672-6	-82	0.0345	-7470A	PPB	28 Sep 2022 11:52:27	U	SMPL	
N4843-01 - 1	HD-SUMPWATER-0926	4136	0.9635	-7470A	PPB	28 Sep 2022 11:54:44	U	SMPL	
N4787-07 - 1	PT-HG-WP	46782	10.3566	-7470A	PPB	28 Sep 2022 11:57:00	U	SMPL	
CCV77 - 1	CCV77	21077	4.6949	-7470A	PPB	28 Sep 2022 12:03:55	U	SMPL	
CCB77 - 1	CCB77	-206	0.0072	-7470A	PPB	28 Sep 2022 12:06:14	U	SMPL	
PB147907TB - 1	PB147907TB	13	0.0554	-7470A	PPB	28 Sep 2022 12:08:31	U	SMPL	
PB147914TB - 1	PB147914TB	30	0.0592	-7470A	PPB	28 Sep 2022 12:10:46	U	SMPL	
N4812-02LX5 - 1		-77	0.0356	-7470A	PPB	28 Sep 2022 12:13:02	U	SMPL	
N4812-02A - 1		17102	3.8194	-7470A	PPB	28 Sep 2022 12:15:18	U	SMPL	
N4821-03LX5 - 1		-164	0.0164	-7470A	PPB	28 Sep 2022 12:17:34	U	SMPL	
N4821-03A - 1		15181	3.3962	-7470A	PPB	28 Sep 2022 12:22:58	U	SMPL	
N4821-01LX5 - 1		-234	0.0010	-7470A	PPB	28 Sep 2022 12:25:13	U	SMPL	
N4821-01A - 1		16873	3.7689	-7470A	PPB	28 Sep 2022 12:27:31	U	SMPL	
N4787-07DLX2 - 1		23333	5.1918	-7470A	PPB	28 Sep 2022 12:29:47	U	SMPL	
CCV78 - 1	CCV78	22204	4.9431	-7470A	PPB	28 Sep 2022 12:32:04	U	SMPL	
CCB78 - 1	CCB78	-174	0.0142	-7470A	PPB	28 Sep 2022 12:34:23	U	SMPL	