

LB122242


7471B

INSTRUMENT ID: CV2

Linear

μ Abs.:

53927



A= 0.0000e+000

B= 1.8485e-004

C= -1.4400e-002

Rho= 0.9998812

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.060	0.060	405	0.000	405					1/00
0.2	0.200	0.177	-0.023	1034	0.0 %	1034					-12
2.5	2.500	2.417	-0.083	13151	0.0 %	13151					-3
5.0	5.000	5.015	0.015	27209	0.0 %	27209					0
7.5	7.500	7.577	0.077	41067	0.0 %	41067					1
10.0	10.000	9.954	-0.046	53927	0.0 %	53927					0

LB122242 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 Sep 2022 14:36:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	405	-	0.0000	7471B	PPB	28 Sep 2022 14:37:19	S	Std
0.2 - 1	50.2	1034	-	0.2000	7471B	PPB	28 Sep 2022 14:39:37	S	Std
2.5 - 1	52.5	13151	-	2.5000	7471B	PPB	28 Sep 2022 14:41:54	S	Std
5.0 - 1	55	27209	-	5.0000	7471B	PPB	28 Sep 2022 14:44:11	S	Std
7.5 - 1	57.5	41067	-	7.5000	7471B	PPB	28 Sep 2022 14:46:30	S	Std
10.0 - 1	59	53927	-	10.0000	7471B	PPB	28 Sep 2022 14:47:48	S	Std
ICV49 - 1	ICV49	21845	4.0237	-	7471B	PPB	28 Sep 2022 14:50:44	U	SMPL
ICB49 - 1	ICB49	294	0.0399	-	7471B	PPB	28 Sep 2022 14:52:59	U	SMPL
CCV80 - 1	CCV80	27042	4.9844	-	7471B	PPB	28 Sep 2022 14:55:16	U	SMPL
CCB80 - 1	CCB80	296	0.0403	-	7471B	PPB	28 Sep 2022 14:57:32	U	SMPL
CRA - 1	CRA	981	0.1669	-	7471B	PPB	28 Sep 2022 14:59:48	U	SMPL
HighStd - 1	HighStd	53783	9.9275	-	7471B	PPB	28 Sep 2022 15:02:03	U	SMPL
ChkStd - 1	ChkStd	38269	7.0597	-	7471B	PPB	28 Sep 2022 15:04:19	U	SMPL
PB147956BL - 1	PBS	313	0.0435	-	7471B	PPB	28 Sep 2022 15:06:38	U	SMPL
PB147956BS - 1	LCSS	21669	3.9911	-	7471B	PPB	28 Sep 2022 15:14:22	U	SMPL
N4803-01 - 1	SASP2-T-1-(10-17)	1473	0.2579	-	7471B	PPB	28 Sep 2022 15:16:37	U	SMPL
N4803-02 - 1	SASP2-T-1-(17-23)	1315	0.2287	-	7471B	PPB	28 Sep 2022 15:18:56	U	SMPL
N4812-01 - 1	WC-3	16354	3.0087	-	7471B	PPB	28 Sep 2022 15:21:13	U	SMPL
N4815-01 - 1	SAND-FILTER-1	37422	6.9031	-	7471B	PPB	28 Sep 2022 15:23:30	U	SMPL
N4815-02 - 1	SAND-FILTER-2	4278	0.7764	-	7471B	PPB	28 Sep 2022 15:25:46	U	SMPL
CCV81 - 1	CCV81	27487	5.0666	-	7471B	PPB	28 Sep 2022 15:28:04	U	SMPL
CCB81 - 1	CCB81	333	0.0472	-	7471B	PPB	28 Sep 2022 15:30:19	U	SMPL
N4826-01 - 1	OK-03-092622	4635	0.8424	-	7471B	PPB	28 Sep 2022 15:32:36	U	SMPL
N4844-01 - 1	SU-04-092622	3208	0.5786	-	7471B	PPB	28 Sep 2022 15:34:51	U	SMPL
N4850-01 - 1	WC-14B-SP-2	188018	34.7410	-	7471B	PPB	28 Sep 2022 15:37:07	U	SMPL
N4855-01 - 1	OE-02-092822	1630	0.2869	-	7471B	PPB	28 Sep 2022 15:39:22	U	SMPL
N4857-01 - 1	OR-02-092722	3350	0.6049	-	7471B	PPB	28 Sep 2022 15:41:48	U	SMPL
N4857-01DUP - 1	OR-02-092722DUP	3344	0.6037	-	7471B	PPB	28 Sep 2022 15:44:04	U	SMPL
N4857-01MS - 1	OR-02-092722MS	27522	5.0731	-	7471B	PPB	28 Sep 2022 15:53:14	U	SMPL
N4857-01MSD - 1	OR-02-092722MSD	25367	4.6747	-	7471B	PPB	28 Sep 2022 15:55:31	U	SMPL
N4857-01LX5 - 1		592	0.0950	-	7471B	PPB	28 Sep 2022 16:00:08	U	SMPL
N4857-01A - 1		24913	4.5908	-	7471B	PPB	28 Sep 2022 16:02:23	U	SMPL
CCV82 - 1	CCV82	27379	5.0467	-	7471B	PPB	28 Sep 2022 16:04:39	U	SMPL
CCB82 - 1	CCB82	347	0.0497	-	7471B	PPB	28 Sep 2022 16:06:57	U	SMPL
N4850-01DLX5 - 1		20598	3.7932	-	7471B	PPB	28 Sep 2022 16:10:15	U	SMPL
CCV83 - 1	CCV83	27472	5.0638	-	7471B	PPB	28 Sep 2022 16:12:31	U	SMPL
CCB83 - 1	CCB83	-114	-0.0355	-	7471B	PPB	28 Sep 2022 16:14:48	U	SMPL