



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB122251

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	10	28.00	82.4	127.4	09/29/2022	09:30
2	N4862-04		1	13	100.00	>212.0	>212.0	09/29/2022	10:00
3	N4862-04DUP		1	12	100.00	>212.0	>212.0	09/29/2022	10:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB/22251

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Flashpoint-0929

WorkList ID : 163627

Department : Wet-Chemistry

Date : 09-29-2022 09:16:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/03/2022	Water	N4862-04	Flash Point	Cool 4 deg C	PSEG03	N11	A2720	09/28/2022	1010B

Date/Time 9/29/2022 9:25
Raw Sample Received by: SJ (WC)
Raw Sample Relinquished by: CP (SM)

Date/Time 9/29/2022 12:00
Raw Sample Received by: CP (SM)
Raw Sample Relinquished by: SJ (WC)