



Analytical Summary Report

Analysis Method: SM 4500-H B

Parameter: pH

Run Number: LB122279

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	09/30/2022	11:05
2	CAL2	1	Water	NA	NA	20.7	7.00	09/30/2022	11:06
3	CAL3	1	Water	NA	NA	20.7	10.03	09/30/2022	11:07
4	ICV	1	Water	NA	NA	20.7	6.99	09/30/2022	11:10
5	CCV1	1	Water	NA	NA	20.7	2.01	09/30/2022	11:12
6	N4788-20	1	Water	NA	NA	22.9	8.99	09/30/2022	11:15
7	N4862-04	1	Water	NA	NA	20.1	7.32	09/30/2022	11:20
8	N4894-01	1	Water	NA	NA	20.1	6.06	09/30/2022	11:24
9	N4894-03	1	Water	NA	NA	20.1	6.83	09/30/2022	11:29
10	N4894-05	1	Water	NA	NA	20.1	6.78	09/30/2022	11:32
11	N4894-07	1	Water	NA	NA	20.2	6.79	09/30/2022	11:35
12	N4894-07DUP	1	Water	NA	NA	20.1	6.80	09/30/2022	11:36
13	CCV2	1	Water	NA	NA	20.7	12.03	09/30/2022	11:40

WORKLIST(Hardcopy Internal Chain)

WB 122279

WorkList Name : ph n4894

WorkList ID : 163676

Department : Wet-Chemistry

Date : 09-30-2022 10:44:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/19/2022	Water	N4788-20	pH	Cool 4 deg C	CHEM02	QA 01	PT-PH-WP	09/19/2022	SM 4500-H B
09/28/2022	Water	N4862-04	pH A	Cool 4 deg C	PSEG03	N11	A2720	09/28/2022	SM 4500-H B
09/28/2022	Water	N4894-01	pH F	Cool 4 deg C	LOCK01	N11	MW-1	09/28/2022	SM 4500-H B
09/28/2022	Water	N4894-03	pH F	Cool 4 deg C	LOCK01	N11	MW-3	09/28/2022	SM 4500-H B
09/28/2022	Water	N4894-05	pH F	Cool 4 deg C	LOCK01	N11	MW-2	09/28/2022	SM 4500-H B
09/28/2022	Water	N4894-07	pH F	Cool 4 deg C	LOCK01	N11	MW-4	09/28/2022	SM 4500-H B

Date/Time 09/30/22 10:50

Raw Sample Received by: **70 (GSC)**

Raw Sample Relinquished by: **[Signature]**

Date/Time 09/30/22 14:00

Raw Sample Received by: **[Signature]**

Raw Sample Relinquished by: **70 (GSC)**