



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB122280

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	09/30/2022	13:05
2	CAL2	1	Water	NA	NA	20.6	7.00	09/30/2022	13:06
3	CAL3	1	Water	NA	NA	20.6	10.03	09/30/2022	13:07
4	ICV	1	Water	NA	NA	20.7	6.99	09/30/2022	13:10
5	CCV1	1	Water	NA	NA	20.6	2.01	09/30/2022	13:12
6	N4788-21	1	Water	NA	NA	22.9	8.99	09/30/2022	13:15
7	N4915-01	1	Water	NA	NA	20.1	8.55	09/30/2022	13:20
8	N4917-01	1	Water	NA	NA	20.1	8.45	09/30/2022	13:22
9	N4919-01	1	Water	NA	NA	20.1	8.46	09/30/2022	13:25
10	N4919-01DUP	1	Water	NA	NA	20.2	8.47	09/30/2022	13:26
11	CCV2	1	Water	NA	NA	20.6	12.03	09/30/2022	13:30

LB 122280

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N4917 WorkList ID : 163679 Department : Wet-Chemistry Date : 09-30-2022 12:54:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/19/2022	Water	N4788-21	pH	Cool 4 deg C	CHEM02	QA Of	PT-PH-WP	09/19/2022	9040C
09/28/2022	Water	N4915-01	pH <u>D</u>	Cool 4 deg C	ENVI70	N11	257155-3	09/28/2022	9040C
09/28/2022	Water	N4917-01	pH <u>D</u>	Cool 4 deg C	ENVI70	N11	265393-5	09/28/2022	9040C
09/28/2022	Water	N4919-01	pH <u>D</u>	Cool 4 deg C	ENVI70	N11	265393-6	09/28/2022	9040C

Date/Time 09/30/22 13:00
Raw Sample Received by: JP (09/)
Raw Sample Relinquished by: JP (09/)

Date/Time 09/30/22 15:00
Raw Sample Received by: JP (09/)
Raw Sample Relinquished by: JP (09/)