



Analytical Summary Report

Analysis Method: 9041A
Parameter: pH
Run Number: LB122281

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : N/A

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940
3.0-5.5 pH paper refill	W1934
Test Paper, PH, Short Range 6.0/8.0	W2350
Test Paper,PH Short Range 9.0/10.0	W1939
0.0-3.0 pH paper refill	W1931
10.0-12.0 pH paper refill	W1940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.0	09/30/2022	13:05
2	CAL2	1	Water	NA	NA	20.6	7.0	09/30/2022	13:06
3	CAL3	1	Water	NA	NA	20.6	10.0	09/30/2022	13:07
4	ICV	1	Water	NA	NA	20.7	6.99	09/30/2022	13:10
5	CCV1	1	Water	NA	NA	20.6	2.0	09/30/2022	13:12
6	N4788-22	1	Water	NA	NA	22.9	9.0	09/30/2022	13:15
7	CCV2	1	Water	NA	NA	20.6	12.0	09/30/2022	13:30

WORKLIST(Hardcopy Internal Chain)

122221

WorkList Name : PH N4788

WorkList ID : 163680

Department : Wet-Chemistry

Date : 09-30-2022 12:55:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/19/2022	Water	N4788-22	pH	Cool 4 deg C	CHEM02	QA OI	PT-PH-WP	09/19/2022	9041A

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

09/30/22 13:00

JP (WOC)

JP (WOC)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

09/30/22

15:00

JP (WOC)

JP (WOC)