



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Jignesh

Run Number: LB122283

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP99252
Turbidity Calibration std, 20NTU	WP99247
Turbidity Calibration std, 40NTU	WP99246
Turbidity Calibration std, 1NTU	WP99250
Turbidity Calibration std, 5NTU	WP99249
Turbidity Calibration - CCV std, 10 NTU	WP99248

Intercept: 0.0861

Slope: 0.9227

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.150	0.07		09/30/2022	11:30
2	CAL2	1	1	1.150	1.15	-15.3	09/30/2022	11:34
3	CAL3	5	1	4.480	4.76	4.8	09/30/2022	11:37
4	CAL4	10	1	9.460	10.16	-1.6	09/30/2022	11:40
5	CAL5	20	1	18.300	19.74	1.3	09/30/2022	11:43
6	CAL6	40	1	37.100	40.12	-0.3	09/30/2022	11:47



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Jignesh

Run Number: LB122283

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.310	9.997	09/30/2022	11:50
2	ICB		1	0.142	0.061	09/30/2022	11:54
3	CCV1	10	1	9.360	10.051	09/30/2022	12:00
4	CCB1		1	0.154	0.074	09/30/2022	12:03
5	LB122283BL		1	0.158	0.078	09/30/2022	12:07
6	N4788-34		1	20.090	21.680	09/30/2022	12:10
7	N4894-01		1	0.368	0.306	09/30/2022	12:13
8	N4894-01DUP		1	0.366	0.303	09/30/2022	12:17
9	N4894-03		1	1.120	1.121	09/30/2022	12:20
10	N4894-05		1	132.000	142.965	09/30/2022	12:23
11	N4894-07		1	157.000	170.059	09/30/2022	12:27
12	CCV2	10	1	9.140	9.812	09/30/2022	12:30
13	CCB2		1	0.148	0.067	09/30/2022	12:33

WORKLIST(Hardcopy Internal Chain)

LB122282

WorkList Name : TURB-093022

WorkList ID : 163674

Department : Wet-Chemistry

Date : 09-30-2022 09:29:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/21/2022	Water	N4788-34	Turbidity	Cool 4 deg C	CHEM02	QA Of	PT-TURB-WP	09/19/2022	SM2130 B
09/30/2022	Water	N4894-01	Turbidity	Cool 4 deg C	LOCK01	N11	MW-1	09/28/2022	SM2130 B
09/30/2022	Water	N4894-03	Turbidity	Cool 4 deg C	LOCK01	N11	MW-3	09/28/2022	SM2130 B
09/30/2022	Water	N4894-05	Turbidity	Cool 4 deg C	LOCK01	N11	MW-2	09/28/2022	SM2130 B
09/30/2022	Water	N4894-07	Turbidity	Cool 4 deg C	LOCK01	N11	MW-4	09/28/2022	SM2130 B

Date/Time 9/30/22 11:00
 Raw Sample Received by: JPC
 Raw Sample Relinquished by: JPC

Date/Time 9/30/22 12:50
 Raw Sample Received by: JPC
 Raw Sample Relinquished by: JPC