



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Jignesh

Run Number: LB122284

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP99252
Turbidity Calibration std, 20NTU	WP99247
Turbidity Calibration std, 40NTU	WP99246
Turbidity Calibration std, 1NTU	WP99250
Turbidity Calibration std, 5NTU	WP99249
Turbidity Calibration - CCV std, 10 NTU	WP99248

Intercept: 0.0861

Slope: 0.9227

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.150	0.08		09/30/2022	11:30
2	CAL2	1	1	1.150	1.16	-16.4	09/30/2022	11:34
3	CAL3	5	1	4.480	4.76	4.7	09/30/2022	11:37
4	CAL4	10	1	9.460	10.15	-1.5	09/30/2022	11:40
5	CAL5	20	1	18.300	19.71	1.5	09/30/2022	11:43
6	CAL6	40	1	37.100	40.14	-0.3	09/30/2022	11:47



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Jignesh

Run Number: LB122284

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.310	9.997	09/30/2022	11:50
2	ICB		1	0.142	0.061	09/30/2022	11:54
3	CCV1	10	1	9.360	10.051	09/30/2022	12:00
4	CCB1		1	0.154	0.074	09/30/2022	12:03
5	LB122284BL		1	0.158	0.078	09/30/2022	12:07
6	N4788-35		1	20.090	21.680	09/30/2022	12:10
7	CCV2	10	1	9.140	9.812	09/30/2022	12:30
8	CCB2		1	0.148	0.067	09/30/2022	12:33