



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Sohil

Run Number: LB122287

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	09/30/2022	16:50
2	CAL2	1	Water	NA	NA	20.6	7.00	09/30/2022	16:51
3	CAL3	1	Water	NA	NA	20.6	10.03	09/30/2022	16:52
4	ICV	1	Water	NA	NA	20.7	7.02	09/30/2022	17:00
5	CCV1	1	Water	NA	NA	20.7	2.01	09/30/2022	17:02
6	N4886-02	1	Solid	20.02	20	21.3	8.66	09/30/2022	17:05
7	N4914-02	1	Solid	20.01	20	21.1	8.24	09/30/2022	17:10
8	N4914-04	1	Solid	20.02	20	21.2	7.59	09/30/2022	17:13
9	N4932-01	1	Solid	20.03	20	20.9	7.76	09/30/2022	17:15
10	N4932-01DUP	1	Solid	20.04	20	20.8	7.77	09/30/2022	17:16
11	CCV2	1	Water	NA	NA	20.7	12.03	09/30/2022	17:20

WORKLIST(Hardcopy Internal Chain)

122287

WorkList Name : CORROSIONITY N4914 S

WorkList ID : 163685

Department : Wet-Chemistry

Date : 09-30-2022 13:06:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/29/2022	Solid	N4886-02	Corrosivity	Cool 4 deg C	PSEG03	N11	OILY-DEBRIS-0929	09/29/2022	9045D
09/29/2022	Solid	N4914-02	Corrosivity	Cool 4 deg C	ENTA05	N11	S1	09/29/2022	9045D
09/29/2022	Solid	N4914-04	Corrosivity	Cool 4 deg C	ENTA05	N11	S2	09/29/2022	9045D
09/30/2022	Solid	N4932-01	Corrosivity	Cool 4 deg C	ICES02	N11	RBR-2520555	09/30/2022	9045D

Date/Time 09/30/22 16:45
 Raw Sample Received by: 70169C,
 Raw Sample Relinquished by: 70169C, (S.M.)

Date/Time 09/30/22 18:00
 Raw Sample Received by: 70169C, (S.M.)
 Raw Sample Relinquished by: 70169C, (S.M.)