



PERCENT SOLID

Supervisor: Sohil

Analyst: JIGNESH

Date: 10/3/2022

OVENTEMP IN Celsius(°C): 108

Time IN: 14:40

In Date: 10/01/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 2.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:00

Out Date: 10/02/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB122292

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4926-01	MH0AC3	1	0.92	9.02	9.94	9.65	96.8	
N4926-02	MH0AC3D	2	0.92	9.02	9.94	9.65	96.8	
N4926-03	MH0AC3S	3	0.92	9.02	9.94	9.65	96.8	
N4926-04	MH0AC4	4	0.93	9.03	9.96	9.52	95.1	
N4926-05	MH0AC5	5	0.9	9.03	9.93	9.41	94.2	
N4926-06	MH0AC6	6	0.91	8.97	9.88	9.15	91.9	
N4926-07	MH0AC7	7	0.89	9.02	9.91	9.12	91.2	
N4926-08	MH0AC8	8	0.91	9.06	9.97	9.33	92.9	
N4926-09	MH0AC9	9	0.92	9.04	9.96	9.21	91.7	
N4926-10	MH0AD0	10	0.9	8.96	9.86	9.63	97.4	
N4926-11	MH0AD1	11	0.91	9.06	9.97	9.38	93.5	
N4926-12	MH0AD2	12	0.89	9.02	9.91	9.42	94.6	
N4926-13	MH0AD3	13	0.93	8.97	9.9	9.48	95.3	
N4926-14	MH0AD4	14	0.89	9.09	9.98	9.87	98.8	
N4926-15	MH0AD5	15	0.9	8.92	9.82	9.31	94.3	
N4926-16	MH0AD6	16	0.88	9.06	9.94	9.41	94.2	
N4926-17	MH0AD7	17	0.9	9.02	9.92	8.77	87.3	
N4926-18	MH0AD8	18	0.87	9.01	9.88	8.97	89.9	
N4926-19	MH0AD9	19	0.89	9.05	9.94	8.77	87.1	
N4926-20	MH0AE0	20	0.91	8.90	9.81	8.85	89.2	
N4926-21	MH0AE1	21	0.89	9.03	9.92	9.08	90.7	
N4926-22	MH0AE2	22	0.87	9.05	9.92	9.3	93.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$