

LB122296


INSTRUMENT ID: CH.

7471B

Linear

 μ Abs.:

52633



A= 0.0000e+000

B= 1.9120e-004

C= 6.2217e-002

Rho= 0.9996798

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.081	0.081	100	0.000	100				
0.2	0.200	0.252	0.052	992	0.0 %	992				
2.5	2.500	2.447	-0.053	12474	0.0 %	12474				
5.0	5.000	4.856	-0.144	25075	0.0 %	25075				
7.5	7.500	7.438	-0.062	38576	0.0 %	38576				
10.0	10.000	10.125	0.125	52633	0.0 %	52633				

LB122296 INSTRUMENT ID:CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 Oct 2022 12:43:37

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	100	-	0.0000	7471B	PPB	02 Oct 2022 12:46:57	S	Std
0.2 - 1	50.2	992	-	0.2000	7471B	PPB	02 Oct 2022 12:49:14	S	Std
2.5 - 1	52.5	12474	-	2.5000	7471B	PPB	02 Oct 2022 12:51:32	S	Std
5.0 - 1	55.0	25075	-	5.0000	7471B	PPB	02 Oct 2022 12:53:49	S	Std
7.5 - 1	57.5	38576	-	7.5000	7471B	PPB	02 Oct 2022 12:56:06	S	Std
10.0 - 1	510.0	52633	-	10.0000	7471B	PPB	02 Oct 2022 12:58:24	S	Std
ICV55 - 1	ICV55	21012	4.0796	-	7471B	PPB	02 Oct 2022 13:04:19	U	SMPL
ICB55 - 1	ICB55	10	0.0641	-	7471B	PPB	02 Oct 2022 13:06:34	U	SMPL
CCV01 - 1	CCV01	27870	5.3908	-	7471B	PPB	02 Oct 2022 13:08:51	U	SMPL
CCB01 - 1	CCB01	-1	0.0620	-	7471B	PPB	02 Oct 2022 13:11:06	U	SMPL
CRA - 1	CRA	991	0.2517	-	7471B	PPB	02 Oct 2022 13:13:24	U	SMPL
HighStd - 1	HighStd	53044	10.2040	-	7471B	PPB	02 Oct 2022 13:27:03	U	SMPL
ChkStd - 1	ChkStd	36857	7.1091	-	7471B	PPB	02 Oct 2022 13:29:19	U	SMPL
PB148045BL - 1	PBS	-61	0.0506	-	7471B	PPB	02 Oct 2022 13:31:38	U	SMPL
PB148045BS - 1	LCSS	20646	4.0096	-	7471B	PPB	02 Oct 2022 13:33:56	U	SMPL
N4862-01 - 1	VNJ-203	5565	1.1262	-	7471B	PPB	02 Oct 2022 13:36:12	U	SMPL
N4868-01 - 1	ROLL-OFF-349-ETGI	3933	0.8142	-	7471B	PPB	02 Oct 2022 13:38:30	U	SMPL
N4892-01 - 1	TOPSOIL-STORMWATER-DENALI	1689	0.3851	-	7471B	PPB	02 Oct 2022 13:40:47	U	SMPL
N4895-01 - 1	CRUSHED-MASONRY-PILE-8	3585	0.7438	-	7471B	PPB	02 Oct 2022 13:43:04	U	SMPL
N4896-01 - 1	CRUSHED-MASONRY-PILE-9	1291	0.3091	-	7471B	PPB	02 Oct 2022 13:45:20	U	SMPL
CCV02 - 1	CCV02	26530	5.1348	-	7471B	PPB	02 Oct 2022 13:47:35	U	SMPL
CCB02 - 1	CCB02	11	0.0643	-	7471B	PPB	02 Oct 2022 13:49:50	U	SMPL
N4897-01 - 1	CRUSHED-MASONRY-PILE-10	372	0.1333	-	7471B	PPB	02 Oct 2022 13:52:07	U	SMPL
N4898-01 - 1	CRUSHED-MASONRY-PILE-11	2307	0.5033	-	7471B	PPB	02 Oct 2022 13:54:22	U	SMPL
N4907-01 - 1	SU-03-092922	812	0.2175	-	7471B	PPB	02 Oct 2022 13:56:38	U	SMPL
N4907-03 - 1	VNJ-222	3323	0.6976	-	7471B	PPB	02 Oct 2022 13:58:53	U	SMPL
N4907-05 - 1	VNJ-232	5557	1.1247	-	7471B	PPB	02 Oct 2022 14:01:09	U	SMPL
N4912-01 - 1	VNJ-211	1217	0.2949	-	7471B	PPB	02 Oct 2022 14:03:25	U	SMPL
N4912-03 - 1	72-11939	726	0.2010	-	7471B	PPB	02 Oct 2022 14:05:41	U	SMPL
N4936-01 - 1	SASP2-T-2-(0-6)	22456	4.3557	-	7471B	PPB	02 Oct 2022 14:07:58	U	SMPL
N4936-01DUP - 1	SASP2-T-2-(0-6)DUP	22208	4.3083	-	7471B	PPB	02 Oct 2022 14:10:16	U	SMPL
N4936-01MS - 1	SASP2-T-2-(0-6)MS	41596	8.0152	-	7471B	PPB	02 Oct 2022 14:12:33	U	SMPL
CCV03 - 1	CCV03	25999	5.0331	-	7471B	PPB	02 Oct 2022 14:14:51	U	SMPL
CCB03 - 1	CCB03	-19	0.0586	-	7471B	PPB	02 Oct 2022 14:17:08	U	SMPL
N4936-01MSD - 1	SASP2-T-2-(0-6)MSD	41651	8.0257	-	7471B	PPB	02 Oct 2022 14:22:26	U	SMPL
N4936-02 - 1	SASP2-T-2-(10-16)	879	0.2303	-	7471B	PPB	02 Oct 2022 14:24:41	U	SMPL
N4936-03 - 1	SASP2-T-2-(16-22)	293	0.1182	-	7471B	PPB	02 Oct 2022 14:27:00	U	SMPL
N4943-01 - 1	SOIL-001-0930	558	0.1689	-	7471B	PPB	02 Oct 2022 14:29:15	U	SMPL
N4943-03 - 1	SOIL-002-0930	5773	1.1660	-	7471B	PPB	02 Oct 2022 14:31:31	U	SMPL
N4943-05 - 1	SOIL-003-0930	11847	2.3273	-	7471B	PPB	02 Oct 2022 14:33:47	U	SMPL
N4946-01 - 1	HD-01-093022	1334	0.3173	-	7471B	PPB	02 Oct 2022 14:36:04	U	SMPL
N4947-01 - 1	CL-02-093022	5935	1.1970	-	7471B	PPB	02 Oct 2022 14:38:21	U	SMPL
CCV04 - 1	CCV04	26278	5.0864	-	7471B	PPB	02 Oct 2022 14:40:39	U	SMPL
CCB04 - 1	CCB04	15	0.0651	-	7471B	PPB	02 Oct 2022 14:42:54	U	SMPL
N4936-01A + OILXS		4452	0.9134	-	7471B	PPB	02 Oct 2022 14:45:12	U	SMPL
N4936-01B + TOIA		41387	7.9752	-	7471B	PPB	02 Oct 2022 14:47:28	U	SMPL
CCV05 - 1	CCV05	26444	5.1182	-	7471B	PPB	02 Oct 2022 14:49:44	U	SMPL
CCB05 - 1	CCB05	17	0.0655	-	7471B	PPB	02 Oct 2022 14:52:02	U	SMPL

8
10/2/22