

Analysis Method: 9050A

ANALYST: Iwona

Parameter: Conductance

SUPERVISOR REVIEW BY: Jignesh

Run Number: LB122319

Meter ID: WC CONDUCTANCE METER-1

Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP98799

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB122319BL		1	20.0	0.83	0.92	10/03/2022	14:10
2	LB122319BS	1413	1	20.1	1280.00	1412.16	10/03/2022	14:13
3	N4788-09		1	20.3	858.00	942.62	10/03/2022	14:16

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$