



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB122323

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID:

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	10	28.00	82.4	82.4	10/04/2022	08:45
2	N4930-01		1	15	100.00	>212.0	>212.0	10/04/2022	09:15
3	N4959-01		1	14	100.00	>212.0	>212.0	10/04/2022	09:47
4	N4961-06		1	12	100.00	>212.0	>212.0	10/04/2022	10:16
5	N4968-01		1	12	100.00	>212.0	>212.0	10/04/2022	10:45
6	N4968-02		1	14	100.00	>212.0	>212.0	10/04/2022	11:16
7	N4968-03		1	13	100.00	>212.0	>212.0	10/04/2022	11:48
8	N4968-04		1	12	100.00	>212.0	>212.0	10/04/2022	12:15
9	N4968-05		1	15	100.00	>212.0	>212.0	10/04/2022	12:47
10	N4970-03		1	14	100.00	>212.0	>212.0	10/04/2022	13:15
11	N4970-03DUP		1	14	100.00	>212.0	>212.0	10/04/2022	13:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 122323

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Flash Point-1004

WorkList ID : 163771

Department : Wet-Chemistry

Date : 10-04-2022 08:33:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/05/2022	Water	N4930-01	Flash Point	Cool 4 deg C	PSEG03	N11	TOTES-0930	09/30/2022	1010B
10/10/2022	Water	N4959-01	Flash Point	Cool 4 deg C	PSEG03	N12	34122	10/03/2022	1010B
10/10/2022	Water	N4961-06	Flash Point	Cool 4 deg C	PSEG03	N12	NWB-1829	10/03/2022	1010B
10/10/2022	Water	N4968-01	Flash Point	Cool 4 deg C	PSEG03	N12	CLI-0120	10/03/2022	1010B
10/10/2022	Water	N4968-02	Flash Point	Cool 4 deg C	PSEG03	N12	CLI-82522	10/03/2022	1010B
10/10/2022	Water	N4968-03	Flash Point	Cool 4 deg C	PSEG03	N12	CLI-0144	10/03/2022	1010B
10/10/2022	Water	N4968-04	Flash Point	Cool 4 deg C	PSEG03	N12	CLI-0620	10/03/2022	1010B
10/10/2022	Water	N4968-05	Flash Point	Cool 4 deg C	PSEG03	N12	CLI-0621	10/03/2022	1010B
10/10/2022	Water	N4970-03	Flash Point	Cool 4 deg C	PSEG03	N12	33837	10/03/2022	1010B

Date/Time

10/4/2022 8:35

Raw Sample Received by:

SJ(WC)

Raw Sample Relinquished by:

CP(SM)

Date/Time

10/4/2022

Raw Sample Received by:

CP(SM)

Raw Sample Relinquished by:

ST(WC)