



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB122324

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	10/04/2022	08:25
2	CAL2	1	Water	NA	NA	20.7	7.00	10/04/2022	08:26
3	CAL3	1	Water	NA	NA	20.7	10.03	10/04/2022	08:29
4	ICV	1	Water	NA	NA	20.5	6.98	10/04/2022	08:33
5	CCV1	1	Water	NA	NA	20.6	2.01	10/04/2022	08:35
6	N4930-01	1	Water	NA	NA	20.7	6.02	10/04/2022	08:40
7	N4959-01	1	Water	NA	NA	20.1	5.48	10/04/2022	08:42
8	N4961-06	1	Water	NA	NA	20.1	7.03	10/04/2022	08:45
9	N4968-01	1	Water	NA	NA	20.1	7.88	10/04/2022	08:50
10	N4968-02	1	Water	NA	NA	20.2	9.10	10/04/2022	09:00
11	N4968-03	1	Water	NA	NA	20.1	8.85	10/04/2022	09:03
12	N4968-04	1	Water	NA	NA	20.6	6.23	10/04/2022	09:07
13	N4968-05	1	Water	NA	NA	20.2	6.06	10/04/2022	09:10
14	N4970-03	1	Water	NA	NA	20.2	6.78	10/04/2022	09:12
15	N4970-03DUP	1	Water	NA	NA	20.3	6.79	10/04/2022	09:13
16	CCV2	1	Water	NA	NA	20.7	12.03	10/04/2022	09:15