



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Iwona

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: jignesh

Run Number: LB122329

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP99270
Residual chlorine ICV-LCS, 0.4PPM	WP99269
Chlorine Calibration std, 0.1ppm	WP99264
Chlorine Calibration std, 0.2ppm	WP99265
Chlorine Calibration std, 0.8ppm	WP99267
Chlorine Calibration std, 0.0ppm	WP99263
Chlorine Calibration std, 1.6ppm	WP99268
Residual Chlorine LOQ 0.1ppm	WP99271
Residual Chlorine Calibration and CCV std, 0	WP99266
Total Chlorine Powder Pillows	W2759

Intercept: 0.0111

Slope: 1.0042

Regression: 0.999863

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	-0.01		10/04/2022	14:20
2	CAL2	0.1	1	0.000	0.110	0.110	0.10	2	10/04/2022	14:23
3	CAL3	0.2	1	0.000	0.220	0.220	0.21	-4	10/04/2022	14:26
4	CAL4	0.4	1	0.000	0.410	0.410	0.40	0.8	10/04/2022	14:30
5	CAL5	0.8	1	0.000	0.830	0.830	0.82	-1.9	10/04/2022	14:33
6	CAL6	1.6	1	0.000	1.610	1.610	1.59	0.5	10/04/2022	14:36

Analysis Method: SM4500 Cl GANALYST: IwonaParameter: Residual ChlorineSUPERVISOR REVIEW BY: jigneshRun Number: LB122329

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4100	0.4100	0.3970	10/04/2022	14:40
2	ICB				1	0.0000	0.0000	0.0000	-0.0110	10/04/2022	14:43
3	CCV1			0.4	1	0.0000	0.4200	0.4200	0.4070	10/04/2022	14:46
4	CCB1				1	0.0000	0.0100	0.0100	-0.0010	10/04/2022	14:50
5	LB122329BL	50	50		1	0.0000	0.0100	0.0100	-0.0010	10/04/2022	14:53
6	LB122329BS	50	50	0.4	1	0.0000	0.4000	0.4000	0.3870	10/04/2022	14:56
7	N4788-19	50	50		1	0.0000	2.2900	2.2900	2.2690	10/04/2022	15:00
8	N4955-07	50	50		1	0.0000	0.0600	0.0600	0.0490	10/04/2022	15:03
9	N4955-08	50	50		1	0.0000	0.1100	0.1100	0.0980	10/04/2022	15:06
10	N4788-19				5	0.0000	0.5000	0.5000	0.4870	10/04/2022	15:10
11	CCV2			0.4	1	0.0000	0.4100	0.4100	0.3970	10/04/2022	15:13
12	CCB2				1	0.0000	0.0000	0.0000	-0.0110	10/04/2022	15:16