



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB122333

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	10/04/2022	16:42
2	CAL2	1	Water	NA	NA	20.7	7.00	10/04/2022	16:43
3	CAL3	1	Water	NA	NA	20.7	10.03	10/04/2022	16:44
4	ICV	1	Water	NA	NA	20.6	7.02	10/04/2022	16:45
5	CCV1	1	Water	NA	NA	20.7	2.01	10/04/2022	16:50
6	N4963-01	1	Water	NA	NA	20.1	8.16	10/04/2022	17:00
7	N4963-01DUP	1	Water	NA	NA	20.2	8.17	10/04/2022	17:01
8	N4966-01	1	Water	NA	NA	20.2	8.35	10/04/2022	17:05
9	N4974-01	1	Water	NA	NA	20.2	8.51	10/04/2022	17:10
10	N4977-01	1	Water	NA	NA	20.1	8.47	10/04/2022	17:11
11	N4979-01	1	Water	NA	NA	20.2	8.38	10/04/2022	17:13
12	N4987-02	1	Water	NA	NA	20.1	8.87	10/04/2022	17:20
13	N4987-03	1	Water	NA	NA	20.1	5.24	10/04/2022	17:24
14	CCV2	1	Water	NA	NA	20.7	12.03	10/04/2022	17:25

WORKLIST(Hardcopy Internal Chain)

122333

WorkList Name : PH N4897 W

WorkList ID : 163787

Department : Wet-Chemistry

Date : 10-04-2022 16:34:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/29/2022	Water	N4963-01	pH	Cool 4 deg C	ENV/70	N11	257155-4	09/29/2022	9040C
09/29/2022	Water	N4966-01	pH	Cool 4 deg C	ENV/70	N11	257248-5	09/29/2022	9040C
09/29/2022	Water	N4974-01	pH	Cool 4 deg C	ENV/70	N11	257248-6	09/29/2022	9040C
09/30/2022	Water	N4977-01	pH	Cool 4 deg C	ENV/70	N11	251032-3	09/30/2022	9040C
09/30/2022	Water	N4979-01	pH	Cool 4 deg C	ENV/70	N11	251032-4	09/30/2022	9040C
10/04/2022	Water	N4987-02	pH	Cool 4 deg C	PSEG03	N11	AUD-1677	10/04/2022	9040C
10/04/2022	Water	N4987-03	pH	Cool 4 deg C	PSEG03	N11	AUD-1678	10/04/2022	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

10/04/22 16:35
 7060C
 7060C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

10/04/22 18:00
 7060C
 7060C