



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB122334

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	10/04/2022	16:42
2	CAL2	1	Water	NA	NA	20.7	7.00	10/04/2022	16:43
3	CAL3	1	Water	NA	NA	20.7	10.03	10/04/2022	16:44
4	ICV	1	Water	NA	NA	20.6	7.02	10/04/2022	16:45
5	CCV1	1	Water	NA	NA	20.6	2.01	10/04/2022	17:30
6	N4988-01	1	Solid	20.03	20	20.2	9.47	10/04/2022	17:33
7	N4988-01DUP	1	Solid	20.04	20	20.1	9.48	10/04/2022	17:34
8	CCV2	1	Water	NA	NA	20.6	12.03	10/04/2022	17:35

WB 122334

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N4988 S

WorkList ID : 163788

Department : Wet-Chemistry

Date : 10-04-2022 16:35:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/04/2022	Solid	N4988-01	pH		Cool 4 deg C	PSEG05 N11	AU-5-100422	10/04/2022	9045D

Date/Time 10/04/22 16:37
Raw Sample Received by: J.P. (sm)
Raw Sample Relinquished by: J.P. (sm)

Date/Time 10/04/22 18:00
Raw Sample Received by: J.P. (sm)
Raw Sample Relinquished by: J.P. (sm)