



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Jignesh

Parameter: Flash Point

Supervisor Review By: Sohil

Run Number: LB122337

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	28.00	82.4	82.7	10/05/2022	09:24
2	N4987-02		1	13	100.00	>212.0	>212.0	10/05/2022	10:05
3	N4987-02DUP		1	14	100.00	>212.0	>212.0	10/05/2022	10:35
4	N4987-03		1	16	100.00	>212.0	>212.0	10/05/2022	11:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 922337
LB 122338

WORKLIST(Hardcopy Internal Chain)

WorkList Name : flash point-1005 WorkList ID : 163798 Department : Wet-Chemistry Date : 10-05-2022 08:59:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/06/2022	Solid	N4969-03	Flash Point	Cool 4 deg C	PSEG03	N12	METER-OIL	10/03/2022	1010B
10/11/2022	Water	N4987-02	Flash Point	Cool 4 deg C	PSEG03	N11	AUD-1677	10/04/2022	1010B
10/11/2022	Water	N4987-03	Flash Point	Cool 4 deg C	PSEG03	N11	AUD-1678	10/04/2022	1010B

Date/Time 10/05/2022 9:00
Raw Sample Received by: SI (WC)
Raw Sample Relinquished by: CP (SM)

Date/Time 10/05/2022 12:00
Raw Sample Received by: CP (SM)
Raw Sample Relinquished by: SI (WC)