

LB 12233Q

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12

Instrument ID : Konelab

10/4/2022 16:47

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.764	0.0	0.069	
ICB1	1.272	0.0	0.001	
CCV1	243.488	0.0	0.176	
CCB1	1.106	0.0	0.001	
PB148106BL	0.898	0.0	0.001	
PB148106BS	96.104	0.0	0.070	
N4788-25	724.692	0.0	0.521	Test limit high
N4788-25DLX2	368.310	0.0	0.265	
CCV2	248.198	0.0	0.179	
CCB2	0.658	0.0	0.001	

N	10
Mean	177.949
SD	231.7279
CV%	130.22

Aquakem v. 7.2AQ1

Results from time period:

Tue Oct 04 16:39:27 2022

Tue Oct 04 16:46:14 2022

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	0.3825	µg/l	10/4/2022 10:05:47	
5.OPPBCN	A	Total CN	P	5.1094	µg/l	10/4/2022 10:05:48	
10PPBCN	A	Total CN	P	9.9553	µg/l	10/4/2022 10:05:49	
50PPBCN	A	Total CN	P	48.8958	µg/l	10/4/2022 10:05:50	
100PPBCN	A	Total CN	P	99.6919	µg/l	10/4/2022 10:05:51	
250PPBCN	A	Total CN	P	251.5864	µg/l	10/4/2022 10:05:52	
500PPBCN	A	Total CN	P	499.3786	µg/l	10/4/2022 10:05:53	
ICV1	S	Total CN	P	94.7644	µg/l	10/4/2022 16:39:27	
ICB1	S	Total CN	P	1.2716	µg/l	10/4/2022 16:39:29	
CCV1	S	Total CN	P	243.4885	µg/l	10/4/2022 16:39:32	
CCB1	S	Total CN	P	1.1059	µg/l	10/4/2022 16:39:33	
PB148106BL	S	Total CN	P	0.8982	µg/l	10/4/2022 16:39:35	
PB148106BS	S	Total CN	P	96.1042	µg/l	10/4/2022 16:46:06	
N4788-25	S	Total CN	P	724.6915	µg/l	10/4/2022 16:46:08	
N4788-25DLX2	S	Total CN	P	368.3097	µg/l	10/4/2022 16:46:09	
CCV2	S	Total CN	P	248.1977	µg/l	10/4/2022 16:46:11	
CCB2	S	Total CN	P	0.6582	µg/l	10/4/2022 16:46:14	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

10/4/2022 10:07


Test Total CN

Accepted 10/4/2022 10:07

Factor 1391
Bias 0.001

Coeff. of det. 0.999979

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	0.3825	0.0000	-
2	5.0PPBCN	0.004	5.1094	5.0000	2.2
3	10PPBCN	0.008	9.9553	10.0000	-0.4
4	50PPBCN	0.036	48.8958	50.0000	-2.2
5	100PPBCN	0.072	99.6919	100.0000	0.3
6	250PPBCN	0.181	251.5864	250.0000	0.6
7	500PPBCN	0.359	499.3786	500.0000	-0.1

12/10/4/22