



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB122344

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	10/05/2022	12:35
2	CAL2	1	Water	NA	NA	20.6	7.00	10/05/2022	12:36
3	CAL3	1	Water	NA	NA	20.6	10.03	10/05/2022	12:37
4	ICV	1	Water	NA	NA	20.7	6.99	10/05/2022	12:39
5	CCV1	1	Water	NA	NA	20.8	2.01	10/05/2022	12:40
6	N4996-01	1	Water	NA	NA	20.1	8.18	10/05/2022	12:42
7	N4998-01	1	Water	NA	NA	20.2	8.19	10/05/2022	12:44
8	N5000-01	1	Water	NA	NA	20.1	9.33	10/05/2022	12:50
9	N5002-01	1	Water	NA	NA	20.1	9.33	10/05/2022	12:51
10	N5009-01	1	Water	NA	NA	20.1	8.44	10/05/2022	13:00
11	N5009-02	1	Water	NA	NA	20.1	8.56	10/05/2022	13:01
12	N5011-01	1	Water	NA	NA	20.1	5.56	10/05/2022	13:05
13	N5011-01DUP	1	Water	NA	NA	20.2	5.57	10/05/2022	13:06
14	CCV2	1	Water	NA	NA	20.9	12.03	10/05/2022	13:10

VB 122344

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n5011 WorkList ID : 163816 Department : Wet-Chemistry Date : 10-05-2022 12:28:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/03/2022	Water	N4996-01	pH C	Cool 4 deg C	ENV170	N11	506D-1	10/03/2022	9040C
10/03/2022	Water	N4998-01	pH C	Cool 4 deg C	ENV170	N11	506D-2	10/03/2022	9040C
10/03/2022	Water	N5000-01	pH C	Cool 4 deg C	ENV170	N11	5486-4	10/03/2022	9040C
10/03/2022	Water	N5002-01	pH C	Cool 4 deg C	ENV170	N11	5486-3	10/03/2022	9040C
10/04/2022	Water	N5009-01	pH C	Cool 4 deg C	ENV170	C11	006-3	10/04/2022	9040C
10/04/2022	Water	N5009-02	pH C	Cool 4 deg C	ENV170	C11	006-4	10/04/2022	9040C
10/03/2022	Water	N5011-01	pH A	Cool 4 deg C	STER14	N11	MINI-CELL	10/03/2022	9040C

Date/Time 10/05/22 12:30
 Raw Sample Received by: JN/WC/
 Raw Sample Relinquished by: CR SM

Date/Time 10/05/22 14:00
 Raw Sample Received by: CR SM
 Raw Sample Relinquished by: JN/WC/