


LB122349 INSTRUMENT ID: CV1

7470A

Linear

 μ Abs.:

4253E



A= 0.0000e+000

B= 2.3344e-004

C= -3.3389e-002

Rho= 0.9988028

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.0	0.000	-0.012	-0.012	90	0.000	90					
0.2	0.200	0.190	-0.010	956	0.0 %	956					
2.5	2.500	2.329	-0.171	10119	0.0 %	10119					
5.0	5.000	5.384	0.384	23208	0.0 %	23208					
7.5	7.500	7.413	-0.087	31898	0.0 %	31898					
10.0	10.000	9.897	-0.103	42538	0.0 %	42538					

LB122349 INSTRUMENT ID:CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Oct 2022 15:26:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1		90	-	0.0000	7470A	PPB	05 Oct 2022 15:31:48	S	Std
0.2 - 1		956	-	0.2000	7470A	PPB	05 Oct 2022 15:34:05	S	Std
2.5 - 1		10119	-	2.5000	7470A	PPB	05 Oct 2022 15:36:23	S	Std
5.0 - 1		23208	-	5.0000	7470A	PPB	05 Oct 2022 15:38:40	S	Std
7.5 - 1		31898	-	7.5000	7470A	PPB	05 Oct 2022 15:40:56	S	Std
10.0 - 1		42538	-	10.0000	7470A	PPB	05 Oct 2022 15:43:14	S	Std
ICV58 - 1	ICV58	17063	3.9498		7470A	PPB	05 Oct 2022 15:51:49	U	SMPL
ICB58 - 1	ICB58	-9	-0.0355		7470A	PPB	05 Oct 2022 15:54:04	U	SMPL
CCV112 - 1 <i>CCV20</i>	CCV112	22113	5.1287		7470A	PPB	05 Oct 2022 15:56:20	U	SMPL
CCB112 - 1 <i>CCB20</i>	CCB112	-27	-0.0397		7470A	PPB	05 Oct 2022 15:58:35	U	SMPL
CRA - 1 <i>9/10/15/22</i>	CRA	923	0.1821		7470A	PPB	05 Oct 2022 16:00:51	U	SMPL
HighStd - 1	HighStd	42859	9.9716		7470A	PPB	05 Oct 2022 16:03:07	U	SMPL
ChkStd - 1	ChkStd	30175	7.0107		7470A	PPB	05 Oct 2022 16:05:22	U	SMPL
PB148137BL - 1	PBW	-60	-0.0474		7470A	PPB	05 Oct 2022 16:07:40	U	SMPL
PB148137BS - 1	LCSW	18434	4.2698		7470A	PPB	05 Oct 2022 16:09:57	U	SMPL
N4957-01 - 1	SB-17	-26	-0.0395		7470A	PPB	05 Oct 2022 16:12:13	U	SMPL
N4957-01DUP - 1	SB-17DUP	54	-0.0208		7470A	PPB	05 Oct 2022 16:14:30	U	SMPL
N4957-01MS - 1	SB-17MS	16142	3.7348		7470A	PPB	05 Oct 2022 16:16:47	U	SMPL
N4957-01MSD - 1	SB-17MSD	15880	3.6736		7470A	PPB	05 Oct 2022 16:19:05	U	SMPL
N4957-02 - 1	SB-18	-18	-0.0376		7470A	PPB	05 Oct 2022 16:21:20	U	SMPL
CCV113 - 1 <i>CCV21</i>	CCV113	21609	5.0110		7470A	PPB	05 Oct 2022 16:23:36	U	SMPL
CCB113 - 1 <i>CCB21</i>	CCB113	-33	-0.0411		7470A	PPB	05 Oct 2022 16:25:51	U	SMPL
N4957-03 - 1 <i>9/10/15/22</i>	SB-19	62	-0.0189		7470A	PPB	05 Oct 2022 16:28:08	U	SMPL
N4957-04 - 1	SB-20	60	-0.0194		7470A	PPB	05 Oct 2022 16:30:23	U	SMPL
N4957-05 - 1	SB-21	67	-0.0177		7470A	PPB	05 Oct 2022 16:32:38	U	SMPL
N4957-06 - 1	SB-22	79	-0.0149		7470A	PPB	05 Oct 2022 16:34:54	U	SMPL
N4957-07 - 1	SB-23	64	-0.0184		7470A	PPB	05 Oct 2022 16:37:09	U	SMPL
N4957-08 - 1	SB-24	66	-0.0180		7470A	PPB	05 Oct 2022 16:39:26	U	SMPL
N4957-09 - 1	SB-25	68	-0.0175		7470A	PPB	05 Oct 2022 16:41:43	U	SMPL
N4957-10 - 1	SB-26	54	-0.0208		7470A	PPB	05 Oct 2022 16:44:00	U	SMPL
N4957-11 - 1	SB-27	69	-0.0173		7470A	PPB	05 Oct 2022 16:46:17	U	SMPL
N4957-12 - 1	SB-28	65	-0.0182		7470A	PPB	05 Oct 2022 16:48:33	U	SMPL
CCV114 - 1 <i>CCV22</i>	CCV114	21350	4.9505		7470A	PPB	05 Oct 2022 16:50:48	U	SMPL
CCB114 - 1 <i>CCB22</i>	CCB114	-35	-0.0416		7470A	PPB	05 Oct 2022 16:53:04	U	SMPL
N4957-13 - 1	SB-29	59	-0.0196		7470A	PPB	05 Oct 2022 16:55:21	U	SMPL
N4957-14 - 1 <i>9/10/15/22</i>	SB-30	55	-0.0205		7470A	PPB	05 Oct 2022 16:57:46	U	SMPL
N4957-15 - 1	SB-31	72	-0.0166		7470A	PPB	05 Oct 2022 17:00:01	U	SMPL
N4969-01 - 1	7823	60	-0.0194		7470A	PPB	05 Oct 2022 17:02:17	U	SMPL
N4957-01LX5 - 1	SB-17L	58	-0.0198		7470A	PPB	05 Oct 2022 17:04:33	U	SMPL
N4957-01A - 1	SB-17A	16476	3.8128		7470A	PPB	05 Oct 2022 17:06:50	U	SMPL
PB148096TB - 1	LEB096	-12	-0.0362		7470A	PPB	05 Oct 2022 17:09:06	U	SMPL
CCV23 - 1	CCV23	21476	4.9800		7470A	PPB	05 Oct 2022 17:11:24	U	SMPL
CCB23 - 1	CCB23	-51	-0.0453		7470A	PPB	05 Oct 2022 17:13:41	U	SMPL