


LB122361 INSTRUMENT ID: CVI

7470A

 μ Abs.:

39482



Linear

A= 0.0000e+000

B= 2.5269e-004

C= -6.4977e-003

Rho= 0.9999282

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	7. D
0.0	0.000	0.008	0.008	56	0.000	56					
0.2	0.200	0.223	0.023	908	0.0 %	908					12
2.5	2.500	2.468	-0.032	9791	0.0 %	9791					-1
5.0	5.000	4.950	-0.050	19614	0.0 %	19614					-1
7.5	7.500	7.582	0.082	30030	0.0 %	30030					1
10.0	10.000	9.970	-0.030	39482	0.0 %	39482					-3

LB122361 INSTRUMENT ID:CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Oct 2022 14:17:32

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	56	-	0.0000	7470A	PPB	06 Oct 2022 14:20:02	S	Std
0.2 - 1	50.2	908	-	0.2000	7470A	PPB	06 Oct 2022 14:22:20	S	Std
2.5 - 1	52.5	9791	-	2.5000	7470A	PPB	06 Oct 2022 14:24:37	S	Std
5.0 - 1	55.0	19614	-	5.0000	7470A	PPB	06 Oct 2022 14:26:55	S	Std
7.5 - 1	57.5	30030	-	7.5000	7470A	PPB	06 Oct 2022 14:29:11	S	Std
10.0 - 1	59.0	39482	-	10.0000	7470A	PPB	06 Oct 2022 14:31:28	S	Std
ICV62 - 1	ICV62	16149	4.0742	-	7470A	PPB	06 Oct 2022 14:40:48	U	SMPL
ICB62 - 1	ICB62	-37	-0.0158	-	7470A	PPB	06 Oct 2022 14:43:03	U	SMPL
CCV31 - 1	CCV31	19021	4.7999	-	7470A	PPB	06 Oct 2022 14:45:19	U	SMPL
CCB31 - 1	CCB31	-53	-0.0199	-	7470A	PPB	06 Oct 2022 14:47:34	U	SMPL
CRA - 1	CRA	847	0.2075	-	7470A	PPB	06 Oct 2022 14:49:50	U	SMPL
HighStd - 1	HighStd	38767	9.7896	-	7470A	PPB	06 Oct 2022 14:52:05	U	SMPL
ChkStd - 1	ChkStd	27575	6.9614	-	7470A	PPB	06 Oct 2022 14:54:21	U	SMPL
PB148143BL - 1	PBW	16349	4.1247	-	7470A	PPB	06 Oct 2022 14:56:39	U	SMPL
PB148143BS - 1	LCSW	-66	-0.0232	-	7470A	PPB	06 Oct 2022 14:58:57	U	SMPL
N4963-01 - 1	257155-4	-52	-0.0196	-	7470A	PPB	06 Oct 2022 15:01:14	U	SMPL
N4966-01 - 1	257248-5	-37	-0.0158	-	7470A	PPB	06 Oct 2022 15:03:31	U	SMPL
N4968-01DLX50 - 1	CLI-0120	38	0.0031	-	7470A	PPB	06 Oct 2022 15:05:48	U	SMPL
N4968-02DLX50 - 1	CLI-82522	142	0.0294	-	7470A	PPB	06 Oct 2022 15:08:04	U	SMPL
N4968-03DLX50 - 1	CLI-0144	22	-0.0009	-	7470A	PPB	06 Oct 2022 15:10:20	U	SMPL
CCV32 - 1	CCV32	18807	4.7459	-	7470A	PPB	06 Oct 2022 15:12:36	U	SMPL
CCB32 - 1	CCB32	-50	-0.0191	-	7470A	PPB	06 Oct 2022 15:14:51	U	SMPL
N4968-05 - 1	CLI-0621	4387	1.1021	-	7470A	PPB	06 Oct 2022 15:17:08	U	SMPL
N4969-02 - 1	7822	-76	-0.0257	-	7470A	PPB	06 Oct 2022 15:19:23	U	SMPL
N4969-02DUP - 1	7822DUP	-1	-0.0068	-	7470A	PPB	06 Oct 2022 15:21:39	U	SMPL
N4969-02MS - 1	7822MS	15362	3.8753	-	7470A	PPB	06 Oct 2022 15:23:55	U	SMPL
N4969-02MSD - 1	7822MSD	15350	3.8723	-	7470A	PPB	06 Oct 2022 15:26:10	U	SMPL
N4974-01 - 1	257248-6	-112	-0.0348	-	7470A	PPB	06 Oct 2022 15:28:28	U	SMPL
N4977-01 - 1	251032-3	-51	-0.0194	-	7470A	PPB	06 Oct 2022 15:30:44	U	SMPL
N4979-01 - 1	251032-4	-35	-0.0153	-	7470A	PPB	06 Oct 2022 15:33:01	U	SMPL
N4987-02DLX5 - 1	AUD-1677	138	0.0284	-	7470A	PPB	06 Oct 2022 15:35:18	U	SMPL
N4996-01 - 1	506D-1	-64	-0.0227	-	7470A	PPB	06 Oct 2022 15:37:34	U	SMPL
CCV33 - 1	CCV33	19179	4.8399	-	7470A	PPB	06 Oct 2022 15:39:50	U	SMPL
CCB33 - 1	CCB33	-67	-0.0234	-	7470A	PPB	06 Oct 2022 15:42:05	U	SMPL
N4998-01 - 1	506D-2	-50	-0.0191	-	7470A	PPB	06 Oct 2022 15:44:22	U	SMPL
N5000-01 - 1	5486-4	-45	-0.0179	-	7470A	PPB	06 Oct 2022 15:46:37	U	SMPL
N5002-01 - 1	5486-3	-45	-0.0179	-	7470A	PPB	06 Oct 2022 15:48:53	U	SMPL
N5009-01 - 1	006-3	-15	-0.0103	-	7470A	PPB	06 Oct 2022 15:51:09	U	SMPL
N5009-02 - 1	006-4	-41	-0.0169	-	7470A	PPB	06 Oct 2022 15:53:26	U	SMPL
N4969-02LX5 - 1	7822L	35	0.0023	-	7470A	PPB	06 Oct 2022 15:55:48	U	SMPL
N4969-02A - 1	7822A	16283	4.1081	-	7470A	PPB	06 Oct 2022 15:58:05	U	SMPL
CCV34 - 1	CCV34	18699	4.7186	-	7470A	PPB	06 Oct 2022 16:00:27	U	SMPL
CCB34 - 1	CCB34	-61	-0.0219	-	7470A	PPB	06 Oct 2022 16:02:47	U	SMPL
N4968-01 - 1	CLI-0120	70	0.0112	-	7470A	PPB	06 Oct 2022 16:07:50	U	SMPL
N4968-02 - 1	CLI-82522	438	0.1042	-	7470A	PPB	06 Oct 2022 16:10:06	U	SMPL
N4968-03 - 1	CLI-0144	329	0.0766	-	7470A	PPB	06 Oct 2022 16:12:22	U	SMPL
N4987-02 - 1	AUD-1677	-120	-0.0368	-	7470A	PPB	06 Oct 2022 16:14:38	U	SMPL
CCV35 - 1	CCV35	19359	4.8853	-	7470A	PPB	06 Oct 2022 16:16:53	U	SMPL
CCB35 - 1	CCB35	-64	-0.0227	-	7470A	PPB	06 Oct 2022 16:19:09	U	SMPL