



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB122365

Analyst By : JIGNESH
Supervisor Review By : Iwona
Slope : 99.2
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

True Value of CCV4 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	10/06/2022	16:00
2	CAL2	1	Water	NA	NA	20.7	7.00	10/06/2022	16:01
3	CAL3	1	Water	NA	NA	20.6	10.03	10/06/2022	16:02
4	ICV	1	Water	NA	NA	20.7	6.98	10/06/2022	16:05
5	CCV1	1	Water	NA	NA	20.7	2.01	10/06/2022	16:10
6	N5008-05	1	Water	NA	NA	20.6	7.65	10/06/2022	16:11
7	N5008-05DUP	1	Water	NA	NA	20.7	7.66	10/06/2022	16:12
8	N5027-01	1	Water	NA	NA	20.1	7.57	10/06/2022	16:15
9	N5027-02	1	Water	NA	NA	20.2	8.07	10/06/2022	16:20
10	N5027-03	1	Water	NA	NA	20.2	8.03	10/06/2022	16:22
11	N5027-04	1	Water	NA	NA	20.5	7.99	10/06/2022	16:25
12	N5027-05	1	Water	NA	NA	20.3	7.89	10/06/2022	16:26
13	N5027-06	1	Water	NA	NA	20.2	10.45	10/06/2022	16:30
14	N5027-07	1	Water	NA	NA	20.2	9.42	10/06/2022	16:32
15	N5027-08	1	Water	NA	NA	20.5	9.26	10/06/2022	16:35
16	CCV2	1	Water	NA	NA	20.7	12.03	10/06/2022	16:40
17	N5027-09	1	Water	NA	NA	20.5	8.86	10/06/2022	16:42
18	N5031-01	1	Water	NA	NA	20.2	8.16	10/06/2022	16:45
19	N5031-02	1	Water	NA	NA	20.6	7.94	10/06/2022	16:48
20	N5033-01	1	Water	NA	NA	20.7	8.41	10/06/2022	16:50

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	N5033-02	1	Water	NA	NA	20.2	8.17	10/06/2022	16:52
22	N5042-02	1	Water	NA	NA	20.2	5.27	10/06/2022	17:00
23	N5042-03	1	Water	NA	NA	20.4	5.44	10/06/2022	17:02
24	N5042-04	1	Water	NA	NA	20.2	6.80	10/06/2022	17:05
25	N5042-05	1	Water	NA	NA	20.1	5.84	10/06/2022	17:10
26	N5042-06	1	Water	NA	NA	20.1	6.83	10/06/2022	17:13
27	CCV3	1	Water	NA	NA	20.7	2.01	10/06/2022	17:15
28	N5042-07	1	Water	NA	NA	20.5	6.26	10/06/2022	17:20
29	N5042-08	1	Water	NA	NA	20.8	6.64	10/06/2022	17:22
30	N5042-08DUP	1	Water	NA	NA	20.9	6.65	10/06/2022	17:23
31	CCV4	1	Water	NA	NA	20.8	12.03	10/06/2022	17:25

122365

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N5039

WorkList ID : 163875

Department : Wet-Chemistry

Date : 10-06-2022 15:41:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/05/2022	Water	N5008-05	pH	Cool 4 deg C	PSEG03	C11	COMP-3	10/05/2022	9040C
10/04/2022	Water	N5027-01	pH	Cool 4 deg C	ATCE02	C11	9-X-22-USED-Z-GREEN	10/04/2022	9040C
10/04/2022	Water	N5027-02	pH	Cool 4 deg C	ATCE02	C11	9-X-23-USED-Z-GREEN	10/04/2022	9040C
10/04/2022	Water	N5027-03	pH	Cool 4 deg C	ATCE02	C11	9-X-24-USED-Z-GREEN	10/04/2022	9040C
10/04/2022	Water	N5027-04	pH	Cool 4 deg C	ATCE02	C11	9-X-25-USED-Z-GREEN	10/04/2022	9040C
10/04/2022	Water	N5027-05	pH	Cool 4 deg C	ATCE02	C11	9-X-26-USED-Z-GREEN	10/04/2022	9040C
10/04/2022	Water	N5027-06	pH	Cool 4 deg C	ATCE02	C11	9-X-27-USED-Z-GREEN	10/04/2022	9040C
10/04/2022	Water	N5027-07	pH	Cool 4 deg C	ATCE02	C11	9-X-28-USED-Z-GREEN	10/04/2022	9040C
10/04/2022	Water	N5027-08	pH	Cool 4 deg C	ATCE02	C11	9-X-29-USED-Z-GREEN	10/04/2022	9040C
10/04/2022	Water	N5027-09	pH	Cool 4 deg C	ATCE02	C11	9-X-30-USED-Z-GREEN	10/04/2022	9040C
10/05/2022	Water	N5031-01	pH	Cool 4 deg C	ENV70	N11	5869-3	10/05/2022	9040C
10/05/2022	Water	N5031-02	pH	Cool 4 deg C	ENV70	N11	5869-4	10/05/2022	9040C
10/05/2022	Water	N5033-01	pH	Cool 4 deg C	ENV70	C11	596C-1	10/05/2022	9040C
10/05/2022	Water	N5033-02	pH	Cool 4 deg C	ENV70	C11	596C-2	10/05/2022	9040C
10/06/2022	Water	N5042-02	pH	Cool 4 deg C	PSEG03	C11	EOR-0614	10/06/2022	9040C
10/06/2022	Water	N5042-03	pH	Cool 4 deg C	PSEG03	C11	EOR-0615	10/06/2022	9040C
10/06/2022	Water	N5042-04	pH	Cool 4 deg C	PSEG03	C11	EOR-0617	10/06/2022	9040C
10/06/2022	Water	N5042-05	pH	Cool 4 deg C	PSEG03	C11	EOR-0126	10/06/2022	9040C
10/06/2022	Water	N5042-06	pH	Cool 4 deg C	PSEG03	C11	EOR-0627	10/06/2022	9040C
10/06/2022	Water	N5042-07	pH	Cool 4 deg C	PSEG03	C11	EOR-0632	10/06/2022	9040C
10/06/2022	Water	N5042-08	pH	Cool 4 deg C	PSEG03	C11	EOR-0631	10/06/2022	9040C

Date/Time 10/06/22 15:50

Raw Sample Received by: 70(WC)

Raw Sample Relinquished by: 70(WC)

Date/Time 10/06/22

Raw Sample Received by: 70(WC)

Raw Sample Relinquished by: 70(WC)