

Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: JIGNESH

Run Number: LB122369

BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP99228
calibration std. phosphate 0.5 ppm	WP99227
calibration std. phosphate 0.3 ppm	WP99226
calibration std. phosphate 0.1 ppm	WP99225
calibration std. phosphate 0.05 ppm	WP99224
calibration std. 0 ppm	WP99223
phosphate CCV std.	WP99229
Combined reagent	WP99325
Phenolphthalein indicator	WP96861
Sodium hydroxide, 1N	WP98644
Phosphosphate LOQ std, 0.05PPM	WP99307
Phosphate ICV-LCS Std	WP99230
Phosphate LOD-MDL Std 0.025ppm	WP99306

Intercept: -0.0007

Slope: 0.6503

Regression: 0.999955

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	0.001		10/06/2022	13:30
2	CAL2	0.05	1	50	50	0.034	0.053	-6	10/06/2022	13:30
3	CAL3	0.10	1	50	50	0.065	0.101	-1	10/06/2022	13:31
4	CAL4	0.30	1	50	50	0.190	0.293	2.3	10/06/2022	13:31
5	CAL5	0.50	1	50	50	0.324	0.499	0.2	10/06/2022	13:32
6	CAL6	1.00	1	50	50	0.651	1.002	-0.2	10/06/2022	13:32



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: JIGNESH

Run Number: LB122369

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.328	0.505	10/06/2022	13:33
2	ICB		1	50	50	0.001	0.003	10/06/2022	13:33
3	CCV1	0.50	1	50	50	0.324	0.499	10/06/2022	14:00
4	CCB1		1	50	50	0.002	0.004	10/06/2022	14:00
5	PB148140BL		1	1.00	50	0.002	0.004	10/06/2022	14:01
6	PB148140BS	0.50	1	1.00	50	0.323	0.498	10/06/2022	14:01
7	N4955-01		1	1.00	50	0.170	0.262	10/06/2022	14:02
8	N4955-02		1	1.00	50	0.032	0.050	10/06/2022	14:02
9	CCV2	0.50	1	50	50	0.327	0.504	10/06/2022	14:03
10	CCB2		1	50	50	0.003	0.006	10/06/2022	14:03