


LB122376 INSTRUMENT ID: CVI

SFAM01.1



Linear

A= 0.0000e+000

B= 2.4034e-004

C= -8.1257e-003

Rho= 0.9998898

Accept=Accepted

slope
y-intercept
9/10/6/22

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	0.000	0.000	35	0.000	35				
0.05	0.050					185	0			
0.20	0.200	0.215	0.015	930	0.0 %	930				
2.50	2.500	2.427	-0.073	10130	0.0 %	10130				
5.00	5.000	5.043	0.043	21018	0.0 %	21018				
7.50	7.500	7.579	0.079	31569	0.0 %	31569				
10.0	10.000	9.935	-0.065	41371	0.0 %	41371				

7/10
8.0
-3
-1
-1
10/6/22

LB122376 INSTRUMENT ID: CV1

Method: SFAM01.1

Operator: Admin

Date of Analysis: 27-Sep-2022 13:31:11 6 OCT 2022 18:48 PM

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.00 - 1	50.0	35	-	0.0000	SFAM01.1	PPB	06 Oct 2022 17:08:34	S	Std
0.20 - 1	50.2	930	-	0.2000	SFAM01.1	PPB	06 Oct 2022 17:10:52	S	Std
2.50 - 1	52.5	10130	-	2.5000	SFAM01.1	PPB	06 Oct 2022 17:13:09	S	Std
5.00 - 1	55.0	21018	-	5.0000	SFAM01.1	PPB	06 Oct 2022 17:15:26	S	Std
7.50 - 1	57.5	31569	-	7.5000	SFAM01.1	PPB	06 Oct 2022 17:17:43	S	Std
10.0 - 1	510.0	41371	-	10.0000	SFAM01.1	PPB	06 Oct 2022 17:19:59	S	Std
ICV63 - 1	ICV63	16022	3.8427	-	SFAM01.1	PPB	06 Oct 2022 17:24:28	U	SMPL
ICB63 - 1	ICB63	-52	-0.0206	-	SFAM01.1	PPB	06 Oct 2022 17:26:43	U	SMPL
CCV035 - 1	CCV035	20369	4.8874	-	SFAM01.1	PPB	06 Oct 2022 17:28:59	U	SMPL
CCB035 - 1	CCB035	-46	-0.0192	-	SFAM01.1	PPB	06 Oct 2022 17:31:14	U	SMPL
PB147908BL - 1	PBW908	32	-0.0004	-	SFAM01.1	PPB	06 Oct 2022 17:33:31	U	SMPL
N4624-01 - 1	MCB8F7	386	0.0846	-	SFAM01.1	PPB	06 Oct 2022 17:35:46	U	SMPL
N4624-02 - 1	MCB8G5	72	0.0092	-	SFAM01.1	PPB	06 Oct 2022 17:38:02	U	SMPL
N4624-03 - 1	MCB8G7	287	0.0609	-	SFAM01.1	PPB	06 Oct 2022 17:40:17	U	SMPL
N4624-04 - 1	MCB8G9	86	0.0125	-	SFAM01.1	PPB	06 Oct 2022 17:42:33	U	SMPL
N4624-05 - 1	MCB8H2	6	-0.0067	-	SFAM01.1	PPB	06 Oct 2022 17:44:49	U	SMPL
N4624-06 - 1	MCB8H4	73	0.0094	-	SFAM01.1	PPB	06 Oct 2022 17:47:06	U	SMPL
N4624-07 - 1	MCB8H8	94	0.0145	-	SFAM01.1	PPB	06 Oct 2022 17:49:23	U	SMPL
N4624-08 - 1	MCB8H9	-20	-0.0129	-	SFAM01.1	PPB	06 Oct 2022 17:51:40	U	SMPL
N4624-09 - 1	MCB8J0	-18	-0.0125	-	SFAM01.1	PPB	06 Oct 2022 17:53:55	U	SMPL
N4624-12 - 1	MCB8J8	2	-0.0076	-	SFAM01.1	PPB	06 Oct 2022 17:56:10	U	SMPL
N4624-13 - 1	MCB8J9	65	0.0075	-	SFAM01.1	PPB	06 Oct 2022 17:58:25	U	SMPL
N4624-14 - 1	MCB8K3	12	-0.0052	-	SFAM01.1	PPB	06 Oct 2022 18:00:41	U	SMPL
PB147910BL - 1	PBW910	29	-0.0012	-	SFAM01.1	PPB	06 Oct 2022 18:02:56	U	SMPL
N4625-01 - 1	MCB8F8	-5	-0.0093	-	SFAM01.1	PPB	06 Oct 2022 18:05:12	U	SMPL
N4625-02 - 1	MCB8G6	80	0.0111	-	SFAM01.1	PPB	06 Oct 2022 18:07:28	U	SMPL
N4625-03 - 1	MCB8G8	18	-0.0038	-	SFAM01.1	PPB	06 Oct 2022 18:09:44	U	SMPL
N4625-04 - 1	MCB8H1	-19	-0.0127	-	SFAM01.1	PPB	06 Oct 2022 18:12:01	U	SMPL
N4625-05 - 1	MCB8H5	-23	-0.0137	-	SFAM01.1	PPB	06 Oct 2022 18:14:17	U	SMPL
N4625-06 - 1	MCB8H6	-20	-0.0129	-	SFAM01.1	PPB	06 Oct 2022 18:16:34	U	SMPL
CCV036 - 1	CCV036	19412	4.6574	-	SFAM01.1	PPB	06 Oct 2022 18:18:51	U	SMPL
CCB036 - 1	CCB036	-15	-0.0117	-	SFAM01.1	PPB	06 Oct 2022 18:21:06	U	SMPL
N4625-07 - 1	MCB8J1	29	-0.0012	-	SFAM01.1	PPB	06 Oct 2022 18:23:23	U	SMPL
N4625-08 - 1	MCB8J2	63	0.0070	-	SFAM01.1	PPB	06 Oct 2022 18:25:38	U	SMPL
N4625-09 - 1	MCB8J3	-34	-0.0163	-	SFAM01.1	PPB	06 Oct 2022 18:27:54	U	SMPL
N4625-10 - 1	MCB8K1	-3	-0.0088	-	SFAM01.1	PPB	06 Oct 2022 18:30:09	U	SMPL
N4625-11 - 1	MCB8K2	23	-0.0026	-	SFAM01.1	PPB	06 Oct 2022 18:32:30	U	SMPL
N4625-12 - 1	MCB8K4	1	-0.0079	-	SFAM01.1	PPB	06 Oct 2022 18:34:46	U	SMPL
CCV037 - 1	CCV037	19950	4.7867	-	SFAM01.1	PPB	06 Oct 2022 18:37:03	U	SMPL
CCB037 - 1	CCB037	-13	-0.0113	-	SFAM01.1	PPB	06 Oct 2022 18:39:19	U	SMPL